



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB98460
Analysis Date: 10/09/2018
BalanceID: WC SC-3
OvenID: EX OVEN-2

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 10/09/2018
Extraction IN Time: 08:00
Extraction OUT Time: 08:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB98460BL	LB98460BL	WATER	2	1000	100	3.0266	3.0266	3.02	3.0266	3.0266	0.0000	0
2	LB98460BS	LB98460BS	WATER	2	1000	100	3.1598	3.1598	3.01	3.1783	3.1783	0.0185	18.5
3	LB98460BSD	LB98460BSD	WATER	2	1000	100	3.1675	3.1675	3.03	3.1862	3.1862	0.0187	18.7
4	J5324-03	OILY-WATER	WATER	2	1000	100	3.1191	3.1191	18.02	3.4373	3.4373	0.3182	318.2
5	J5326-42	FL-PH-2	WATER	2	1000	100	2.9226	2.9226	18.04	3.3694	3.3694	0.4468	446.8
6	J5326-45	FL-PH-3	WATER	2	1000	100	3.2064	3.2064	3.05	3.2152	3.2152	0.0088	8.8



QC Batch# LB98460 Test: TPH Analysis Date: 10/09/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2419
pH Paper 0-14	W2416
Sodium Sulfate	EP2519
1:1 HCL	WP64228
Silica Gel	W2368
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP64229
LCSWD	10.00 ML	WP64230
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>10:11</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>09:10</u>		
Bal Check Time:	<u>08:20</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:45</u>
		Out Time:	<u>10:10</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:51</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>11:15</u>		
Bal Check Time:	<u>13:10</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>12:30</u>
		Out Time:	<u>11:50</u>		

LB 98460

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH J5362

WorkList ID : 117740

Date : 10/9/2018 7:44:42 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2018	Water	J5324-03 - 12	TPH	Conc H2SO4 tc	PSEG01	P13	OILY-WATER	10/08/2018	1664A
10/11/2018	Water	J5326-42 - 12	TPH	Conc H2SO4 tc	PSEG01	P21	FL-PH-2	10/08/2018	1664A
10/11/2018	Water	J5326-45 - 12	TPH	Conc H2SO4 tc	PSEG01	P21	FL-PH-3	10/08/2018	1664A

Date/Time 10-09-18 7:50 AM
Received by: JQ
Relinquished by: (6)

Date/Time 10-09-18 10:10 AM
Received by: [Signature]
Relinquished by: JP