



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB98464

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.9	4.01	10/08/2018	17:15
2	CAL2	1	Water			23.0	7.00	10/08/2018	17:17
3	CAL3	1	Water			22.8	10.04	10/08/2018	17:19
4	ICV	1	Water			22.9	7.01	10/08/2018	17:21
5	CCV1	1	Water			22.8	2.00	10/08/2018	17:55
6	J5324-04	1	Water			21.7	7.27	10/08/2018	17:57
7	J5324-04DUP	1	Water			21.8	7.28	10/08/2018	17:59
8	J5326-43	1	Water			22.5	6.56	10/08/2018	18:01
9	J5326-46	1	Water			22.2	7.30	10/08/2018	18:03
10	CCV2	1	Water			22.7	12.00	10/08/2018	18:05

WORKLIST(Hardcopy Internal Chain)

LB 98464

WorkList Name : corrosivity100818-1

WorkList ID : 117727

Date : 10/8/2018 4:55:09 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/08/2018	Water	J5324-04	Corrosivity	Cool 4 deg C	PSEG01	P13	OILY-WATER	10/08/2018	9040C
10/08/2018	Water	J5326-43	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PH-2	10/08/2018	9040C
10/08/2018	Water	J5326-46	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PH-3	10/08/2018	9040C

Date/Time 10/08/18 5:30

Received by: SP

Relinquished by: CP

Date/Time 10/08/18 6:10

Received by: CP

Relinquished by: SP