



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98475

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	10/09/2018	10:20
2	J5324-04		1	9	>100.0	>212.0	>212.0	10/09/2018	10:50
3	J5324-04DUP		1	9	>100.0	>212.0	>212.0	10/09/2018	11:15
4	J5326-43		1	8	>100.0	>212.0	>212.0	10/09/2018	11:40
5	J5326-46		1	7	>100.0	>212.0	>212.0	10/09/2018	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB98475

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp10918

WorkList ID : 117768

Date : 10/9/2018 9:45:22 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2018	Water	J5324-04	Flash Point	Cool 4 deg C	PSEG01	P13	OILY-WATER	10/08/2018	1010A
10/11/2018	Water	J5326-43	Flash Point	Cool 4 deg C	PSEG01	P21	FL-PH-2	10/08/2018	1010A
10/11/2018	Water	J5326-46	Flash Point	Cool 4 deg C	PSEG01	P21	FL-PH-3	10/08/2018	1010A

Date/Time 10/9/18 9:50 AM
Received by: KS
Relinquished by: KS

Date/Time 10/9/18 12:15 PM
Received by: KS
Relinquished by: KS