

LB98486

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

10/9/2018 15:40

Test: *Reactive*
~~Total~~ CN *mg/L*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.177	0.0	0.089	
ICB1	0.940	0.0	0.004	
CCV1	245.007	0.0	0.221	
CCB1	0.606	0.0	0.003	
PB113681BL	0.287	0.0	0.003	
J5324-04	0.186	0.0	0.003	
J5324-04DUP	0.335	0.0	0.003	
J5326-43	0.315	0.0	0.003	
J5326-46	0.412	0.0	0.003	
CCV2	248.302	0.0	0.224	
CCB2	0.857	0.0	0.004	

N	11
Mean	54.039
SD	99.5018
CV%	184.13

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/9/2018 15:45


Test Total CN

Accepted 10/9/2018 12:59

Factor 1124
Bias 0.003

Coeff. of det. 0.999992

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.4172	0.0000	
2	5.0PPBCN	0.007	5.2722	5.0000	
3	10PPBCN	0.012	10.4917	10.0000	
4	50PPBCN	0.047	49.6519	50.0000	
5	100PPBCN	0.091	99.1599	100.0000	
6	250PPBCN	0.225	249.6337	250.0000	
7	500PPBCN	0.448	500.3734	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Tue Oct 09 15:40:38 2018

Tue Oct 09 15:40:48 2018

AS

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	0.4172	µg/l	10/9/2018 12:53:26	
5.0PPBCN	A	Reactive CN	P	5.2722	µg/l	10/9/2018 12:53:27	
10PPBCN	A	Reactive CN	P	10.4917	µg/l	10/9/2018 12:53:28	
50PPBCN	A	Reactive CN	P	49.6519	µg/l	10/9/2018 12:53:29	
100PPBCN	A	Reactive CN	P	99.1599	µg/l	10/9/2018 12:53:30	
250PPBCN	A	Reactive CN	P	249.6337	µg/l	10/9/2018 12:53:31	
500PPBCN	A	Reactive CN	P	500.3734	µg/l	10/9/2018 12:53:32	
ICV1	S	Reactive CN	P	97.1767	µg/l	10/9/2018 15:40:38	
ICB1	S	Reactive CN	P	0.94	µg/l	10/9/2018 15:40:39	
CCV1	S	Reactive CN	P	245.0072	µg/l	10/9/2018 15:40:40	
CCB1	S	Reactive CN	P	0.6064	µg/l	10/9/2018 15:40:41	
PB113681BL	S	Reactive CN	P	0.2872	µg/l	10/9/2018 15:40:42	
J5324-04	S	Reactive CN	P	0.1857	µg/l	10/9/2018 15:40:43	
J5324-04DUP	S	Reactive CN	P	0.3353	µg/l	10/9/2018 15:40:44	
J5326-43	S	Reactive CN	P	0.3153	µg/l	10/9/2018 15:40:45	
J5326-46	S	Reactive CN	P	0.4124	µg/l	10/9/2018 15:40:46	
CCV2	S	Reactive CN	P	248.3015	µg/l	10/9/2018 15:40:47	
CCB2	S	Reactive CN	P	0.8569	µg/l	10/9/2018 15:40:48	