



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/10/2018

OVENTEMP IN Celsius(°C): 107
Time IN: 16:54
In Date: 10/09/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:47
Out Date: 10/10/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB98487

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J5321-01	MGACK5	1	1.13	9.71	9.35	95.8
J5321-02	MGACK6	2	1.14	9.63	9.31	96.2
J5321-03	MGACK7	3	1.13	9.51	9.04	94.4
J5321-04	MGACK8	4	1.13	9.66	9.19	94.5
J5321-05	MGACK9	5	1.18	9.96	9.69	96.9
J5321-06	MGACL0	6	1.12	9.54	8.59	88.7
J5321-07	MGACL0D	7	1.12	9.54	8.59	88.7
J5321-08	MGACL0S	8	1.12	9.54	8.59	88.7
J5321-09	MGACL1	9	1.19	9.54	9.06	94.3
J5321-10	MGACL2	10	1.16	9.57	9.08	94.2
J5321-11	MGACL3	11	1.11	9.53	9.36	98.0
J5321-12	MGACL4	12	1.18	9.61	9.36	97.0
J5321-13	MGACL5	13	1.19	9.79	9.51	96.7
J5321-14	MGACL6	14	1.15	9.54	8.73	90.3
J5321-15	MGACL7	15	1.14	9.85	9.43	95.2
J5321-16	MGACL8	16	1.18	9.95	9.17	91.1
J5321-17	MGACL9	17	1.12	9.5	9.08	95.0
J5321-18	MGACM0	18	1.17	9.63	9.1	93.7
J5321-19	MGACM1	19	1.17	9.96	9.21	91.5
J5321-20	MGACM2	20	1.17	9.67	9.45	97.4
J5321-21	MGACM3	21	1.18	9.75	9.31	94.9
J5321-22	MGACM4	22	1.13	9.66	8.58	87.3
J5321-23	MGACM5	23	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$