



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB98492

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.7	4.01	10/09/2018	16:40
2	CAL2	1	Water			22.8	7.00	10/09/2018	16:42
3	CAL3	1	Water			22.7	10.04	10/09/2018	16:44
4	ICV	1	Water			22.6	7.01	10/09/2018	16:46
5	CCV1	1	Water			22.7	2.00	10/09/2018	16:48
6	J5374-01	1	Solid	20.02	20	23.3	5.06	10/09/2018	16:50
7	J5374-01DUP	1	Solid	20.03	20	23.4	5.07	10/09/2018	16:52
8	J5374-02	1	Solid	20.03	20	23.1	5.86	10/09/2018	16:54
9	J5374-03	1	Solid	20.02	20	23.1	6.18	10/09/2018	16:56
10	CCV2	1	Water			22.7	12.00	10/09/2018	16:58

LB980092

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH100918-2

WorkList ID : 117803

Date : 10/9/2018 4:24:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/03/2018	Solid	J5374-01	pH	Cool 4 deg C	CTIA01	P23	WJ-18-10-09-TSHG-15	10/03/2018	9045D
10/03/2018	Solid	J5374-02	pH	Cool 4 deg C	CTIA01	P23	WJ-18-10-09-TSHG-16	10/03/2018	9045D
10/03/2018	Solid	J5374-03	pH	Cool 4 deg C	CTIA01	P23	WJ-18-10-09-TSHG-17	10/03/2018	9045D

Date/Time 10/09/18 4:30
 Received by: SP
 Relinquished by: CP

Date/Time 10/09/18 5:00
 Received by: CP
 Relinquished by: SP