



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/10/2018

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 10/09/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:21
Out Date: 10/10/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB98493

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J5343-01	380927724.0	1	1.14	9.52	1.25	1.3
J5344-01	380920023.0	2	1.17	9.83	2.62	16.7
J5345-01	380920013.0	3	1.17	9.98	1.49	3.6
J5346-01	380920128.1	4	1.1	9.97	1.33	2.6
J5347-01	380920039.1	5	1.14	9.73	1.37	2.7
J5351-01	380917108.0	6	1.1	9.98	1.31	2.4
J5352-01	380926723.0	7	1.12	9.72	1.35	2.7
J5353-01	380913024.0	8	1.11	9.92	1.74	7.2
J5354-01	380920037.0	9	1.14	9.97	3.18	23.1
J5355-01	380919760.0	10	1.18	9.67	3.25	24.4
J5366-01	RBR200035	11	1.19	9.73	7.64	75.5
J5366-02	RBR200035	12	1.19	9.73	7.64	75.5
J5369-01	CONCRETE-PILE	13	1.15	9.6	8.94	92.2
J5369-02	CONCRETE-PILE-A	14	1.15	9.6	8.94	92.2
J5369-03	CONCRETE-PILE-B	15	1.15	9.6	8.94	92.2
J5369-04	CONCRETE-PILE-C	16	1.15	9.6	8.94	92.2
J5369-05	CONCRETE-PILE-D	17	1.15	9.69	8.94	91.2
J5369-06	CONCRETE-PILE-E	18	1.15	9.69	8.94	91.2
J5369-07	CONCRETE-PILE-F	19	1.15	9.69	8.94	91.2
J5369-08	CONCRETE-PILE-G	20	1.15	9.69	8.94	91.2
J5369-09	CONCRETE-PILE-H	21	1.15	9.69	8.94	91.2
J5372-01	OIL	22	1.00	2.00	2.00	100.0
J5372-02	OILY-DEBRI	23	1.00	2.00	2.00	100.0
J5373-01	RB37708-7211990-45911RT	24	1.18	9.98	8.7	85.5
J5373-03	RT1499-3783-RT3026	25	1.16	9.76	8.52	85.6
J5373-08	SV253553LC	26	1.00	2.00	2.00	100.0
J5374-01	WJ-18-10-09-TSHG-15	27	1.1	9.82	7.31	71.2
J5374-02	WJ-18-10-09-TSHG-16	28	1.18	9.98	7.78	75.0
J5374-03	WJ-18-10-09-TSHG-17	29	1.15	9.78	7.06	68.5

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$