

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/10/2018

TEMP1 IN: 103 10/10/2018 07:00 TEMP OUT: 104 10/10/2018 08:10

TEMP2 IN: 104 10/10/2018 09:00 TEMP OUT: 104 10/10/2018 10:05

TEMP3 IN: 105 10/10/2018 11:30 TEMP OUT: 105 10/10/2018 13:05

TEMP4 IN: 105 10/10/2018 14:00 TEMP OUT: 105 10/10/2018 15:35

Run Number: LB98499

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98499BL	LB98499BL	1.3890	1.3890	100	1.3890	1.3890	1.3890	0.0000	0
2	LB98499BS	LB98499BS	1.4042	1.4042	100	1.4550	1.4550	1.4550	0.0508	508
3	J5371-01	100818-JETT-F-D-B	1.3977	1.3977	200	1.4298	1.4298	1.4298	0.0321	160.5
4	J5371-01DUP	100818-JETT-F-D-BDUP	1.4059	1.4059	200	1.4375	1.4375	1.4375	0.0316	158
5	J5371-02	100818-JETT-F-D-M	1.4010	1.4010	200	1.4290	1.4290	1.4290	0.0280	140
6	J5371-03	100818-JETT-F-D-S	1.3935	1.3935	200	1.4113	1.4113	1.4113	0.0178	89
7	J5371-04	100818-JETT-F-U-B	1.3920	1.3920	100	1.4259	1.4259	1.4259	0.0339	339
8	J5371-05	100818-JETT-F-U-M	1.3927	1.3927	200	1.4227	1.4227	1.4227	0.0300	150
9	J5371-06	100818-JETT-F-U-S	1.3974	1.3974	300	1.4151	1.4151	1.4151	0.0177	59
10	J5371-07	100818-TIPC-F-D-B	1.4057	1.4057	300	1.4218	1.4218	1.4218	0.0161	53.7
11	J5371-08	100818-TIPC-F-D-S	1.4066	1.4066	300	1.4201	1.4201	1.4201	0.0135	45
12	J5371-09	100818-TIPC-F-U-B	1.4058	1.4058	300	1.4191	1.4191	1.4191	0.0133	44.3
13	J5371-10	100818-TIPC-F-U-S	1.4009	1.4009	300	1.4128	1.4128	1.4128	0.0119	39.7
14	J5371-11	100818-SIDI-E-D-B	1.3859	1.3859	300	1.4043	1.4043	1.4043	0.0184	61.3
15	J5371-12	100818-SIDI-E-D-M	1.4104	1.4104	300	1.4203	1.4203	1.4203	0.0099	33

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/10/2018

TEMP1 IN: 103 10/10/2018 07:00 TEMP OUT: 104 10/10/2018 08:10

TEMP2 IN: 104 10/10/2018 09:00 TEMP OUT: 104 10/10/2018 10:05

TEMP3 IN: 105 10/10/2018 11:30 TEMP OUT: 105 10/10/2018 13:05

TEMP4 IN: 105 10/10/2018 14:00 TEMP OUT: 105 10/10/2018 15:35

Run Number: LB98499

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5371-13	100818-SIDI-E-D-S	1.3973	1.3973	400	1.4099	1.4099	1.4099	0.0126	31.5
17	J5371-14	100818-SIDI-E-U-B	1.4006	1.4006	200	1.4265	1.4265	1.4265	0.0259	129.5
18	J5371-15	100818-SIDI-E-U-M	1.4046	1.4046	300	1.4272	1.4272	1.4272	0.0226	75.3
19	J5371-16	100818-SIDI-E-U-S	1.4005	1.4005	200	1.4044	1.4044	1.4044	0.0039	19.5
20	J5371-16DUP	100818-SIDI-E-U-SDUP	1.4012	1.4012	200	1.4050	1.4050	1.4050	0.0038	19

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS101018

WorkList ID : 117815

Date : 10/10/2018 8:03:04 AM

1398499

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2018	Water	J5371-01	TSS	Cool 4 deg C	HDR103	P13	100818-JETT-F-D-B	10/08/2018	SM2540 D
10/15/2018	Water	J5371-02	TSS	Cool 4 deg C	HDR103	P13	100818-JETT-F-D-M	10/08/2018	SM2540 D
10/15/2018	Water	J5371-03	TSS	Cool 4 deg C	HDR103	P13	100818-JETT-F-D-S	10/08/2018	SM2540 D
10/15/2018	Water	J5371-04	TSS	Cool 4 deg C	HDR103	P13	100818-JETT-F-U-B	10/08/2018	SM2540 D
10/15/2018	Water	J5371-05	TSS	Cool 4 deg C	HDR103	P13	100818-JETT-F-U-M	10/08/2018	SM2540 D
10/15/2018	Water	J5371-06	TSS	Cool 4 deg C	HDR103	P13	100818-JETT-F-U-S	10/08/2018	SM2540 D
10/11/2018	Water	J5371-07	TSS	Cool 4 deg C	HDR103	P13	100818-TIPC-F-D-B	10/08/2018	SM2540 D
10/11/2018	Water	J5371-08	TSS	Cool 4 deg C	HDR103	P13	100818-TIPC-F-D-S	10/08/2018	SM2540 D
10/11/2018	Water	J5371-09	TSS	Cool 4 deg C	HDR103	P13	100818-TIPC-F-U-B	10/08/2018	SM2540 D
10/11/2018	Water	J5371-10	TSS	Cool 4 deg C	HDR103	P13	100818-TIPC-F-U-S	10/08/2018	SM2540 D
10/15/2018	Water	J5371-11	TSS	Cool 4 deg C	HDR103	P13	100818-SIDI-E-D-B	10/08/2018	SM2540 D
10/15/2018	Water	J5371-12	TSS	Cool 4 deg C	HDR103	P13	100818-SIDI-E-D-M	10/08/2018	SM2540 D
10/15/2018	Water	J5371-13	TSS	Cool 4 deg C	HDR103	P13	100818-SIDI-E-D-S	10/08/2018	SM2540 D
10/15/2018	Water	J5371-14	TSS	Cool 4 deg C	HDR103	P13	100818-SIDI-E-U-B	10/08/2018	SM2540 D
10/15/2018	Water	J5371-15	TSS	Cool 4 deg C	HDR103	P13	100818-SIDI-E-U-M	10/08/2018	SM2540 D
10/15/2018	Water	J5371-16	TSS	Cool 4 deg C	HDR103	P13	100818-SIDI-E-U-S	10/08/2018	SM2540 D

FP-9823888

Date/Time

Received by:

Relinquished by:

10/10/18, 08:30
AS
CP

Date/Time

Received by:

Relinquished by:

10/10/18, 12:30
AS
AS