


7471B

LB 98501
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42470



A= 0.0000e+000

B= 2.3699e-004

C= -1.1777e-002

Rho= 0.9995888

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	62	0.000	62				
0.2	0.200	0.186	-0.014	835	0.0 %	835				
2.5	2.500	2.656	0.156	11256	0.0 %	11256				
5.0	5.000	4.800	-0.200	20305	0.0 %	20305				
7.5	7.500	7.502	0.002	31704	0.0 %	31704				
10.0	10.000	10.053	0.053	42470	0.0 %	42470				

LB98501 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Oct 2018 08:01:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	62	-	0.00007471B	PPB		10 Oct 2018 08:06:16	S	Std
0.2 - 1	SO.2	835	-	0.20007471B	PPB		10 Oct 2018 08:11:05	S	Std
2.5 - 1	SO.5	11256	-	2.50007471B	PPB		10 Oct 2018 08:13:23	S	Std
5.0 - 1	SO	20305	-	5.00007471B	PPB		10 Oct 2018 08:15:40	S	Std
7.5 - 1	SO.75	31704	-	7.50007471B	PPB		10 Oct 2018 08:23:16	S	Std
10.0 - 1	SO	42470	-	10.00007471B	PPB		10 Oct 2018 08:25:32	S	Std
ICV15 - 1	ICV15	16122	3.8090	-7471B	PPB		10 Oct 2018 08:31:04	U	SMPL
ICB15 - 1	ICB15	127	0.0183	-7471B	PPB		10 Oct 2018 08:33:19	U	SMPL
CCV42 - 1	CCV42	21066	4.9806	-7471B	PPB		10 Oct 2018 08:38:33	U	SMPL
CCB42 - 1	CCB42	157	0.0254	-7471B	PPB		10 Oct 2018 08:40:49	U	SMPL
CRA - 1	CRA	798	0.1773	-7471B	PPB		10 Oct 2018 08:43:06	U	SMPL
HighStd - 1	HighStd	42002	9.9423	-7471B	PPB		10 Oct 2018 08:45:21	U	SMPL
ChkStd - 1	ChkStd	29580	6.9984	-7471B	PPB		10 Oct 2018 08:47:38	U	SMPL
PB113690BL - 1	PBS	177	0.0302	-7471B	PPB		10 Oct 2018 08:49:57	U	SMPL
PB113690BS - 1	LCSS	14558	3.4383	-7471B	PPB		10 Oct 2018 08:52:16	U	SMPL
J5366-01 - 1	RBR200035	3296	0.7693	-7471B	PPB		10 Oct 2018 08:54:33	U	SMPL
J5366-01DUP - 1	RBR200035DUP	2848	0.6632	-7471B	PPB		10 Oct 2018 08:59:34	U	SMPL
J5366-01MS - 1	RBR200035MS	18197	4.3007	-7471B	PPB		10 Oct 2018 09:01:52	U	SMPL
J5366-01MSD - 1	RBR200035MSD	17842	4.2166	-7471B	PPB		10 Oct 2018 09:04:10	U	SMPL
J5369-01 - 1	CONCRETE-PILE	1067	0.2411	-7471B	PPB		10 Oct 2018 09:06:28	U	SMPL
CCV43 - 1	CCV43	20778	4.9124	-7471B	PPB		10 Oct 2018 09:08:45	U	SMPL
CCB43 - 1	CCB43	37	-0.0030	-7471B	PPB		10 Oct 2018 09:11:02	U	SMPL
J5373-01 - 1	RB37708-7211990-45911RT	7282	1.7140	-7471B	PPB		10 Oct 2018 09:13:19	U	SMPL
J5373-03 - 1	RT1499-3783-RT3026	14954	3.5322	-7471B	PPB		10 Oct 2018 09:15:35	U	SMPL
J5374-01 - 1	WJ-18-10-09-TSHG-15	2416	0.5608	-7471B	PPB		10 Oct 2018 09:17:51	U	SMPL
J5374-02 - 1	WJ-18-10-09-TSHG-16	5401	1.2682	-7471B	PPB		10 Oct 2018 09:20:08	U	SMPL
J5374-03 - 1	WJ-18-10-09-TSHG-17	2222	0.5148	-7471B	PPB		10 Oct 2018 09:22:25	U	SMPL
J5366-01LX5 - 1		739	0.1634	-7471B	PPB		10 Oct 2018 09:24:42	U	SMPL
J5366-01A - 1		20451	4.8349	-7471B	PPB		10 Oct 2018 09:27:00	U	SMPL
CCV44 - 1	CCV44	20611	4.8728	-7471B	PPB		10 Oct 2018 09:29:17	U	SMPL
CCB44 - 1	CCB44	130	0.0190	-7471B	PPB		10 Oct 2018 09:31:36	U	SMPL