



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB98503

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.7	4.01	10/09/2018	16:40
2	CAL2	1	Water			22.8	7.00	10/09/2018	16:42
3	CAL3	1	Water			22.7	10.04	10/09/2018	16:44
4	ICV	1	Water			22.6	7.01	10/09/2018	16:46
5	CCV1	1	Water			22.5	2.00	10/09/2018	17:10
6	J5365-02	1	Water			22.5	5.06	10/09/2018	17:12
7	J5365-02DUP	1	Water			22.6	5.07	10/09/2018	17:14
8	J5367-02	1	Water			22.2	6.87	10/09/2018	17:16
9	CCV2	1	Water			22.6	12.00	10/09/2018	17:18

WORKLIST(Hardcopy Internal Chain)

LB98503

WorkList Name : CORROSIONITY100918-2

WorkList ID : 117792

Date : 10/9/2018 2:53:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/09/2018	Water	J5365-02 ☒	Corrosivity	Cool 4 deg C	PSEG01	P21	CY-1	10/09/2018	9040C
10/09/2018	Water	J5367-02 ☒	Corrosivity	Cool 4 deg C	PSEG01	P21	DS-VB-16202	10/09/2018	9040C

Date/Time 10/09/18 5:00
 Received by: SP
 Relinquished by: CP

Date/Time 10/09/18 5:40
 Received by: CP
 Relinquished by: SP