


LB98505
INSTRUMENT ID: CV4

7470A

μ Abs.:

41637



Linear

A= 0.0000e+000

B= 2.4037e-004

C= 2.1946e-002

Rho= 0.9993343

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	-127	0.000	-127				
0.2	0.200	0.219	0.019	819	0.0 %	819				
2.5	2.500	2.398	-0.102	9884	0.0 %	9884				
5.0	5.000	5.249	0.249	21746	0.0 %	21746				
7.5	7.500	7.313	-0.187	30333	0.0 %	30333				
10.0	10.000	10.030	0.030	41637	0.0 %	41637				

LB98505 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Oct 2018 10:17:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-127	-	0.00007470A	PPB		10 Oct 2018 10:40:04	S	Std
0.2 - 1	50.2	819	-	0.20007470A	PPB		10 Oct 2018 10:42:12	S	Std
2.5 - 1	52.5	9884	-	2.50007470A	PPB		10 Oct 2018 10:44:21	S	Std
5.0 - 1	55	21746	-	5.00007470A	PPB		10 Oct 2018 10:46:30	S	Std
7.5 - 1	57.5	30333	-	7.50007470A	PPB		10 Oct 2018 10:48:38	S	Std
10.0 - 1	57.0	41637	-	10.00007470A	PPB		10 Oct 2018 10:53:17	S	Std
ICV16 - 1	ICV16	16551	4.0003	-7470A	PPB		10 Oct 2018 11:06:21	U	SMPL
ICB16 - 1	ICB16	-232	-0.0338	-7470A	PPB		10 Oct 2018 11:08:27	U	SMPL
CCV45 - 1	CCV45	21113	5.0968	-7470A	PPB		10 Oct 2018 11:10:34	U	SMPL
CCB45 - 1	CCB45	-153	-0.0148	-7470A	PPB		10 Oct 2018 11:12:41	U	SMPL
CRA - 1	CRA	896	0.2373	-7470A	PPB		10 Oct 2018 11:17:11	U	SMPL
HighStd - 1	HighStd	41565	10.0128	-7470A	PPB		10 Oct 2018 11:19:17	U	SMPL
ChkStd - 1	ChkStd	27742	6.6902	-7470A	PPB		10 Oct 2018 11:21:23	U	SMPL
PB113671BL - 1	PBW	-322	-0.0555	-7470A	PPB		10 Oct 2018 11:23:33	U	SMPL
PB113671BS - 1	LCSW	14151	3.4234	-7470A	PPB		10 Oct 2018 11:25:43	U	SMPL
PB113671TB - 1	PB113644TB	-180	-0.0213	-7470A	PPB		10 Oct 2018 11:27:51	U	SMPL
J5200-06 - 1	OR-03-100818-A	-96	-0.0011	-7470A	PPB		10 Oct 2018 11:29:59	U	SMPL
J5200-06DUP - 1	OR-03-100818-ADUP	17	0.0260	-7470A	PPB		10 Oct 2018 11:32:08	U	SMPL
J5200-06MS - 1	OR-03-100818-AMS	14622	3.5366	-7470A	PPB		10 Oct 2018 11:34:17	U	SMPL
J5200-06MSD - 1	OR-03-100818-AMSD	14532	3.5150	-7470A	PPB		10 Oct 2018 11:36:24	U	SMPL
CCV46 - 1	CCV46	20637	4.9824	-7470A	PPB		10 Oct 2018 11:38:30	U	SMPL
CCB46 - 1	CCB46	-285	-0.0466	-7470A	PPB		10 Oct 2018 11:40:38	U	SMPL
J5200-08 - 1	OR-03-100818-B	-47	0.0106	-7470A	PPB		10 Oct 2018 11:42:46	U	SMPL
J5201-02 - 1	EO-02-100818-A	-45	0.0111	-7470A	PPB		10 Oct 2018 11:44:53	U	SMPL
J5201-04 - 1	EO-02-100818-B	-35	0.0135	-7470A	PPB		10 Oct 2018 11:46:59	U	SMPL
J5201-06 - 1	EO-02-100818-C	-5	0.0207	-7470A	PPB		10 Oct 2018 11:49:06	U	SMPL
J5318-02 - 1	CON-220-1	21	0.0270	-7470A	PPB		10 Oct 2018 11:51:14	U	SMPL
J5318-04 - 1	CON-220-6	15	0.0256	-7470A	PPB		10 Oct 2018 11:53:21	U	SMPL
J5324-02 - 1	OILY-DEBRIS	6	0.0234	-7470A	PPB		10 Oct 2018 11:55:29	U	SMPL
J5325-02 - 1	SOIL-DEBRI	-48	0.0104	-7470A	PPB		10 Oct 2018 11:57:37	U	SMPL
J5329-04 - 1	POND-1	8	0.0239	-7470A	PPB		10 Oct 2018 11:59:47	U	SMPL
J5329-08 - 1	POND-2	5	0.0231	-7470A	PPB		10 Oct 2018 12:01:54	U	SMPL
CCV47 - 1	CCV47	20649	4.9853	-7470A	PPB		10 Oct 2018 12:04:00	U	SMPL
CCB47 - 1	CCB47	-264	-0.0415	-7470A	PPB		10 Oct 2018 12:06:07	U	SMPL
J5331-03 - 1	NM-STONE-DRUM-3	11	0.0246	-7470A	PPB		10 Oct 2018 12:08:16	U	SMPL
PB113676BL - 1	PBW	-17	0.0179	-7470A	PPB		10 Oct 2018 12:10:23	U	SMPL
PB113676BS - 1	LCSW	14700	3.5553	-7470A	PPB		10 Oct 2018 12:12:30	U	SMPL
PB113676TB - 1	PB113672TB	-179	-0.0211	-7470A	PPB		10 Oct 2018 12:14:37	U	SMPL
J5324-04 - 1	OILY-WATER	33	0.0299	-7470A	PPB		10 Oct 2018 12:16:44	U	SMPL
J5326-43 - 1	FL-PH-2	-49	0.0102	-7470A	PPB		10 Oct 2018 12:18:52	U	SMPL
J5326-43DUP - 1	FL-PH-2DUP	4	0.0229	-7470A	PPB		10 Oct 2018 12:21:00	U	SMPL
J5326-43MS - 1	FL-PH-2MS	14283	3.4551	-7470A	PPB		10 Oct 2018 12:23:08	U	SMPL
J5326-43MSD - 1	FL-PH-2MSD	14578	3.5260	-7470A	PPB		10 Oct 2018 12:25:16	U	SMPL
J5326-46 - 1	FL-PH-3	-122	-0.0074	-7470A	PPB		10 Oct 2018 12:27:23	U	SMPL
CCV48 - 1	CCV48	21144	5.1043	-7470A	PPB		10 Oct 2018 12:29:32	U	SMPL
CCB48 - 1	CCB48	-240	-0.0357	-7470A	PPB		10 Oct 2018 12:31:39	U	SMPL
PB113675BL - 1	PBW	-36	0.0133	-7470A	PPB		10 Oct 2018 12:33:47	U	SMPL
PB113675BS - 1	LCSW	14848	3.5909	-7470A	PPB		10 Oct 2018 12:35:54	U	SMPL
J5304-01 - 1	CA-AQIDW-20181004	-261	-0.0408	-7470A	PPB		10 Oct 2018 12:38:01	U	SMPL
J5324-03 - 1	OILY-WATER	1587	0.4034	-7470A	PPB		10 Oct 2018 12:40:09	U	SMPL
J5326-42 - 1	FL-PH-2	-72	0.0046	-7470A	PPB		10 Oct 2018 12:42:17	U	SMPL
J5326-45 - 1	FL-PH-3	584	0.1623	-7470A	PPB		10 Oct 2018 12:44:25	U	SMPL
J5326-45DUP - 1	FL-PH-3DUP	550	0.1541	-7470A	PPB		10 Oct 2018 12:46:33	U	SMPL
J5326-45MS - 1	FL-PH-3MS	16265	3.9315	-7470A	PPB		10 Oct 2018 12:48:41	U	SMPL
J5326-45MSD - 1	FL-PH-3MSD	16438	3.9731	-7470A	PPB		10 Oct 2018 12:50:50	U	SMPL
CCV49 - 1	CCV49	21215	5.1213	-7470A	PPB		10 Oct 2018 12:52:57	U	SMPL
CCB49 - 1	CCB49	-221	-0.0312	-7470A	PPB		10 Oct 2018 12:55:05	U	SMPL
J5200-06LX5 - 1		-95	-0.0009	-7470A	PPB		10 Oct 2018 12:57:14	U	SMPL
J5200-06A - 1		14962	3.6183	-7470A	PPB		10 Oct 2018 12:59:21	U	SMPL
J5326-43LX5 - 1		-182	-0.0218	-7470A	PPB		10 Oct 2018 13:01:29	U	SMPL
J5326-43A - 1		15205	3.6767	-7470A	PPB		10 Oct 2018 13:03:37	U	SMPL
J5326-45LX5 - 1		5	0.0231	-7470A	PPB		10 Oct 2018 13:05:44	U	SMPL
J5326-45A - 1		16968	4.1005	-7470A	PPB		10 Oct 2018 13:07:51	U	SMPL
CCV50 - 1	CCV50	21140	5.1033	-7470A	PPB		10 Oct 2018 13:09:59	U	SMPL
CCB50 - 1	CCB50	-204	-0.0271	-7470A	PPB		10 Oct 2018 13:12:09	U	SMPL