



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB98510

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP68595
Residual chlorine std, 0PPM	WP68591
Residual chlorine std, 0.1PPM	WP68592
Residual chlorine LOQ, 0.2PPM	WP68601
Residual chlorine std, 0.2PPM	WP68593
Residual chlorine std, and CCV , 0.4PPM	WP68594
Residual chlorine std, 1.6PPM	WP68596
Residual chlorine ICV-LCS, 0.4PPM	WP68597
Residual chlorine LOD, 0.05PPM	WP68598
Total Chlorine Powder Pillows	W1960

Intercept: -0.0031

Slope: 1.0545

Regression: 0.999954

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	10/10/2018	11:44
2	CAL2	0.1	1	0.000	0.100	0.100	10/10/2018	11:45
3	CAL3	0.2	1	0.000	0.200	0.200	10/10/2018	11:46
4	CAL4	0.4	1	0.000	0.420	0.420	10/10/2018	11:48
5	CAL5	0.8	1	0.000	0.850	0.850	10/10/2018	11:49
6	CAL6	1.6	1	0.000	1.680	1.680	10/10/2018	11:51



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB98510

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.430	0.430	0.412	10/10/2018	11:52
2	ICB				1	0.000	0.000	0.000	0.004	10/10/2018	11:53
3	CCV1			0.4	1	0.000	0.410	0.410	0.393	10/10/2018	11:56
4	CCB1				1	0.000	0.000	0.000	0.004	10/10/2018	11:57
5	LB98510BL	50	50		1	0.000	0.000	0.000	0.004	10/10/2018	11:57
6	LB98510BS	50	50	0.4	1	0.000	0.420	0.420	0.403	10/10/2018	11:58
7	J4880-19	50	50		1	0.000	1.810	1.810	1.723	10/10/2018	11:59
8	J5164-07	50	50		1	0.000	0.070	0.070	0.070	10/10/2018	12:00
9	J5164-08	50	50		1	0.000	0.210	0.210	0.203	10/10/2018	12:02
10	J5164-09	50	50		1	0.000	0.060	0.060	0.061	10/10/2018	12:03
11	J4880-19				2	0.000	0.920	0.920	0.878	10/10/2018	12:10
12	CCV2			0.4	1	0.000	0.410	0.410	0.392	10/10/2018	12:11
13	CCB2				1	0.000	0.000	0.000	0.004	10/10/2018	12:12

WORKLIST(Hardcopy Internal Chain)

WorkList Name : RS101018

WorkList ID : 117852

Date : 10/10/2018 1:59:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/28/2018	Water	J5164-07	Residual Chlorine	Cool 4 deg C	CHEM02	Wet Chem L	LOD-MDL-WATER-01-QT4-20	09/28/2018	SM4500 Cl G
09/28/2018	Water	J5164-08	Residual Chlorine	Cool 4 deg C	CHEM02	Wet Chem L	LOQ-WATER-02-QT4-2018	09/28/2018	SM4500 Cl G
09/28/2018	Water	J5164-09	Residual Chlorine	Cool 4 deg C	CHEM02	Wet Chem L	MDL-WATER-03-QT4-2018	09/28/2018	SM4500 Cl G

Date/Time _____
 Received by: _____
 Relinquished by: _____

Date/Time _____
 Received by: _____
 Relinquished by: _____