

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/10/2018

OVEN TEMP1: 103 °C TEMP: 105 °C

DATE & TIME IN: 10/09/2018 16:00 OUT: 10/10/2018 10:00

OVEN TEMP2: 105 °C TEMP: 105 °C

DATE & TIME IN: 10/10/2018 11:20 OUT: 10/10/2018 13:10

Run Number: LB98518

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB98518BL	LB98518BL	82.5647	82.5647	100	82.5647	82.5647	82.5647	0.0000	0
2	J5330-01	EP1-WW01	84.9423	84.9423	100	85.0690	85.0690	85.0690	0.1267	1267
3	J5330-01DUP	EP1-WW01DUP	77.5862	77.5862	100	77.7160	77.7160	77.7160	0.1298	1298
4	J5360-01	2-AVE-22	85.4543	85.4543	100	85.6456	85.6456	85.6456	0.1913	1913

A = Sample Volume (ml)
 B = Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS100918

WorkList ID : 117786

Date : 10/9/2018 1:40:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/12/2018	Water	J5330-01	TS	Cool 4 deg C	CHAZ02	N21	EP1-WW01	10/05/2018	SM2540 B
10/16/2018	Water	J5360-01	TS	Cool 4 deg C	DIST01	P11	2-AVE-22	10/09/2018	SM2540 B

Date/Time 10/09/18, 12:00
 Received by: 
 Relinquished by: 

Date/Time 10/09/18, 15:00
 Received by: 
 Relinquished by: 