



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 10/11/2018

OVENTEMP IN Celsius(°C): 108  
Time IN: 18:00  
In Date: 10/10/2018  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:10  
Out Date: 10/11/2018  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M Sc-1

QC:LB98529

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| J5388-01 | MYAKN7           | 1     | 1.14              | 9.68                          | 9.04                              | 92.5    |
| J5388-02 | MYAKN8           | 2     | 1.17              | 9.83                          | 9.34                              | 94.3    |
| J5388-03 | MYAKP1           | 3     | 1.17              | 9.6                           | 9.16                              | 94.8    |
| J5388-04 | MYAKP2           | 4     | 1.18              | 9.92                          | 8.77                              | 86.8    |
| J5388-05 | MYAKY9           | 5     | 1.16              | 9.59                          | 8.53                              | 87.4    |
| J5388-06 | MYAJX4           | 6     | 1.15              | 9.85                          | 7.58                              | 73.9    |
| J5388-07 | MYAKN3           | 7     | 1.18              | 9.9                           | 9.61                              | 96.7    |
| J5388-08 | MYAKN9           | 8     | 1.19              | 9.87                          | 9.28                              | 93.2    |
| J5388-09 | MYAJW6           | 9     | 1.19              | 9.76                          | 9.45                              | 96.4    |
| J5388-10 | MYAJW7           | 10    | 1.1               | 9.75                          | 9.39                              | 95.8    |
| J5388-11 | MYAJW8           | 11    | 1.19              | 9.58                          | 9.3                               | 96.7    |
| J5388-12 | MYAJW9           | 12    | 1.17              | 9.53                          | 9.29                              | 97.1    |
| J5388-13 | MYAJX0           | 13    | 1.17              | 9.97                          | 9.39                              | 93.4    |
| J5388-14 | MYAJX1           | 14    | 1.1               | 9.63                          | 8.2                               | 83.2    |
| J5388-15 | MYAJX2           | 15    | 1.19              | 9.6                           | 9.04                              | 93.3    |
| J5388-16 | MYAJX3           | 16    | 1.17              | 9.94                          | 7.56                              | 72.9    |
| J5388-17 | MYAJY1           | 17    | 1.19              | 9.8                           | 7.5                               | 73.3    |
| J5388-18 | MYAJY2           | 18    | 1.12              | 9.96                          | 8.5                               | 83.5    |
| J5388-19 | MYAJY3           | 19    | 1.16              | 9.82                          | 8.31                              | 82.6    |
| J5388-20 | MYAJY4           | 20    | 1.1               | 9.84                          | 9.21                              | 92.8    |
| J5388-21 | MYAJY4D          | 21    | 1.1               | 9.84                          | 9.21                              | 92.8    |
| J5388-22 | MYAJY4S          | 22    | 1.1               | 9.84                          | 9.21                              | 92.8    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

1898529

WorkList Name : %1-J5388

WorkList ID : 117873

Date : 10/10/2018 5:36:45 PM

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|-----------------|--------------|--------------|
| 10/12/2018 | Solid  | J5388-01 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAKN7          | 10/02/2018   | Chemtech -SO |
| 10/12/2018 | Solid  | J5388-02 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAKN8          | 10/02/2018   | Chemtech -SO |
| 10/12/2018 | Solid  | J5388-03 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAKP1          | 10/02/2018   | Chemtech -SO |
| 10/12/2018 | Solid  | J5388-04 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAKP2          | 10/02/2018   | Chemtech -SO |
| 10/11/2018 | Solid  | J5388-05 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAKY9          | 10/01/2018   | Chemtech -SO |
| 10/15/2018 | Solid  | J5388-06 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJX4          | 10/05/2018   | Chemtech -SO |
| 10/14/2018 | Solid  | J5388-07 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAKN3          | 10/04/2018   | Chemtech -SO |
| 10/11/2018 | Solid  | J5388-08 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAKN9          | 10/01/2018   | Chemtech -SO |
| 10/14/2018 | Solid  | J5388-09 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJW6          | 10/04/2018   | Chemtech -SO |
| 10/14/2018 | Solid  | J5388-10 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJW7          | 10/04/2018   | Chemtech -SO |
| 10/14/2018 | Solid  | J5388-11 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJW8          | 10/04/2018   | Chemtech -SO |
| 10/14/2018 | Solid  | J5388-12 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJW9          | 10/04/2018   | Chemtech -SO |
| 10/14/2018 | Solid  | J5388-13 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJX0          | 10/04/2018   | Chemtech -SO |
| 10/14/2018 | Solid  | J5388-14 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJX1          | 10/04/2018   | Chemtech -SO |
| 10/15/2018 | Solid  | J5388-15 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJX2          | 10/05/2018   | Chemtech -SO |
| 10/15/2018 | Solid  | J5388-16 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJX3          | 10/05/2018   | Chemtech -SO |
| 10/14/2018 | Solid  | J5388-17 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJY1          | 10/04/2018   | Chemtech -SO |
| 10/14/2018 | Solid  | J5388-18 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJY2          | 10/04/2018   | Chemtech -SO |
| 10/14/2018 | Solid  | J5388-19 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJY3          | 10/04/2018   | Chemtech -SO |
| 10/13/2018 | Solid  | J5388-20 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJY4          | 10/03/2018   | Chemtech -SO |
| 10/13/2018 | Solid  | J5388-21 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJY4D         | 10/03/2018   | Chemtech -SO |

Date/Time 10-10-18 5:40 PM Date/Time 10-10-18 6:10 PM  
 Received by: JD Received by: CP  
 Relinquished by: \_\_\_\_\_ Relinquished by: JP

1898529

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-J5388

WorkList ID : 117873

Date : 10/10/2018 5:36:45 PM

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|-----------------|--------------|--------------|
| 10/13/2018 | Solid  | J5388-22 | Percent Solids | Cool 4 deg C | USEP01   | C21              | MYAJY4S         | 10/03/2018   | Chemtech -SO |

Date/Time 10-10-18 5:40 PM  
 Received by: Jo  
 Relinquished by: Jo

Date/Time 10-10-18 6:10 PM  
 Received by: Jo  
 Relinquished by: Jo