



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB98535

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	10/10/2018	16:40
2	CAL2	1	Water			22.5	7.00	10/10/2018	16:42
3	CAL3	1	Water			22.7	10.03	10/10/2018	16:44
4	ICV	1	Water			22.6	7.01	10/10/2018	16:46
5	CCV1	1	Water			22.5	2.00	10/10/2018	16:48
6	J5271-01	1	Water			18.4	7.42	10/10/2018	16:50
7	J5271-01DUP	1	Water			18.5	7.43	10/10/2018	16:52
8	CCV2	1	Water			22.6	12.00	10/10/2018	16:54

WORKLIST(Hardcopy Internal Chain) LB 98535

WorkList Name : pH101018

WorkList ID : 117867

Date : 10/10/2018 4:26:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/10/2018	Water	J5271-01	pH	Cool 4 deg C	WAYN02	N43	WASTEWATER	10/10/2018	9040C

Date/Time 10/10/18 4:30
 Received by: SP
 Relinquished by: CP

Date/Time 10/10/18 5:00
 Received by: CP
 Relinquished by: SP