



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB98536

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	10/10/2018	16:40
2	CAL2	1	Water			22.5	7.00	10/10/2018	16:42
3	CAL3	1	Water			22.7	10.03	10/10/2018	16:44
4	ICV	1	Water			22.6	7.01	10/10/2018	16:46
5	CCV1	1	Water			22.5	2.00	10/10/2018	17:02
6	J5392-02	1	Solid	20.02	20	23.6	8.34	10/10/2018	17:04
7	J5392-02DUP	1	Solid	20.03	20	23.7	8.35	10/10/2018	17:06
8	CCV2	1	Water			22.7	12.00	10/10/2018	17:08

B98536

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity101018 WorkList ID : 117866 Date : 10/10/2018 4:20:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/10/2018	Solid	J5392-02	Corrosivity	Cool 4 deg C	PSEG01	P23	SOIL-DRUMS	10/10/2018	9045D

Date/Time 10/10/18 4:30
Received by: SP
Relinquished by: CP

Date/Time 10/10/18 5:15
Received by: CP
Relinquished by: SP