

ISM02.4


LB98537

INSTRUMENT ID: CY2

Linear

 μ Abs.:

42514



A= 0.0000e+000

B= 2.3378e-004

C= -2.9347e-002

Rho= 0.9998742

Accept=Accepted

slope
y = mx + c

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.018	-0.018	48	0.000	48					1
0.2	0.200	0.185	-0.015	919	0.0 %	919					1
2.5	2.500	2.499	-0.001	10815	0.0 %	10815					0
5.0	5.000	5.017	0.017	21584	0.0 %	21584					0
7.5	7.500	7.608	0.108	32667	0.0 %	32667					1
10.0	10.000	9.910	-0.090	42514	0.0 %	42514					1

NS
10/11/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	48	-		0 ISM02.4	PPB	10/11/2018 8:58	Std
	0.2 S0.2	919	-		0.2 ISM02.4	PPB	10/11/2018 9:00	Std
	2.5 S2.5	10815	-		2.5 ISM02.4	PPB	10/11/2018 9:03	Std
	5 S5	21584	-		5 ISM02.4	PPB	10/11/2018 9:05	Std
	7.5 S7.5	32667	-		7.5 ISM02.4	PPB	10/11/2018 9:13	Std
	10 S10.0	42514	-		10 ISM02.4	PPB	10/11/2018 9:18	Std
ICV	ICV	17965	4.1705 -		ISM02.4	PPB	10/11/2018 9:20	SMPL
ICB	ICB	-133	-0.0604 -		ISM02.4	PPB	10/11/2018 9:23	SMPL
CCV051	CCV051	21994	5.1124 -		ISM02.4	PPB	10/11/2018 9:25	SMPL
CCB051	CCB051	-21	-0.0343 -		ISM02.4	PPB	10/11/2018 9:27	SMPL
PB113735BL	PBS004	118	-0.0018 -		ISM02.4	PPB	10/11/2018 9:30	SMPL
J5226-04	MA3YY0	251	0.0293 -		ISM02.4	PPB	10/11/2018 9:34	SMPL
J5226-05	MA3YZ2	671	0.1275 -		ISM02.4	PPB	10/11/2018 9:36	SMPL
J5226-06	MA3Z04	1079	0.2229 -		ISM02.4	PPB	10/11/2018 9:39	SMPL
J5226-07	MA3Z04D	1061	0.2187 -		ISM02.4	PPB	10/11/2018 9:41	SMPL
J5226-08	MA3Z04S	12621	2.9212 -		ISM02.4	PPB	10/11/2018 9:43	SMPL
J5226-09	MA3Z05	97	-0.0067 -		ISM02.4	PPB	10/11/2018 9:46	SMPL
J5226-10	MA3Z06	41	-0.0198 -		ISM02.4	PPB	10/11/2018 9:48	SMPL
J5226-11	MA3Z07	686	0.131 -		ISM02.4	PPB	10/11/2018 9:50	SMPL
J5226-12	MA3Z09	215	0.0209 -		ISM02.4	PPB	10/11/2018 9:52	SMPL
J5226-13	MA3Z10	1188	0.2484 -		ISM02.4	PPB	10/11/2018 9:55	SMPL
J5226-14	MA3Z13	906	0.1825 -		ISM02.4	PPB	10/11/2018 9:57	SMPL
J5226-15	MA3Z15	-20	-0.034 -		ISM02.4	PPB	10/11/2018 9:59	SMPL
PB113736BL	PBS004	26	-0.0233 -		ISM02.4	PPB	10/11/2018 10:02	SMPL
J5297-01	MA3YY3	4699	1.0692 -		ISM02.4	PPB	10/11/2018 10:09	SMPL
J5297-02	MA3YY3D	4619	1.0505 -		ISM02.4	PPB	10/11/2018 10:11	SMPL
J5297-03	MA3YY3S	15809	3.6665 -		ISM02.4	PPB	10/11/2018 10:13	SMPL
J5297-04	MA3YZ6	-6	-0.0307 -		ISM02.4	PPB	10/11/2018 10:16	SMPL
J5297-05	MA3Z02	172	0.0109 -		ISM02.4	PPB	10/11/2018 10:18	SMPL
CCV052	CCV052	21856	5.0801 -		ISM02.4	PPB	10/11/2018 10:20	SMPL
CCB052	CCB052	-70	-0.0457 -		ISM02.4	PPB	10/11/2018 10:23	SMPL
J5297-08	MA3Z48	994	0.203 -		ISM02.4	PPB	10/11/2018 10:27	SMPL
J5297-10	MA3YZ7	500	0.0875 -		ISM02.4	PPB	10/11/2018 10:29	SMPL
J5297-11	MA3YZ8	245	0.0279 -		ISM02.4	PPB	10/11/2018 10:32	SMPL
J5297-12	MA3YZ9	225	0.0233 -		ISM02.4	PPB	10/11/2018 10:34	SMPL
J5297-13	MA3Z00	-27	-0.0357 -		ISM02.4	PPB	10/11/2018 10:36	SMPL
J5297-14	MA3Z01	67	-0.0137 -		ISM02.4	PPB	10/11/2018 10:39	SMPL
J5297-15	MA3Z47	614	0.1142 -		ISM02.4	PPB	10/11/2018 10:41	SMPL
J5226-02	MA3YX6	37960	8.8449 -		ISM02.4	PPB	10/11/2018 10:43	SMPL
J5297-06	MA3Z45	9535	2.1997 -		ISM02.4	PPB	10/11/2018 10:50	SMPL
CCV053	CCV053	21358	4.9637 -		ISM02.4	PPB	10/11/2018 10:58	SMPL
CCB053	CCB053	-20	-0.034 -		ISM02.4	PPB	10/11/2018 11:00	SMPL