

7470A


LB98544

INSTRUMENT ID: CY2

Linear

 μ Abs.:

36206



A= 0.0000e+000

B= 2.7562e-004

C= -1.4329e-002

Rho= 0.9999700

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	33	0.000	33				
0.2	0.200	0.211	0.011	816	0.0 %	816				
2.5	2.500	2.476	-0.024	9035	0.0 %	9035				
5.0	5.000	5.000	-0.000	18192	0.0 %	18192				
7.5	7.500	7.554	0.054	27460	0.0 %	27460				
10.0	10.000	9.965	-0.035	36206	0.0 %	36206				

LB98544 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Oct 2018 11:34:21

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	33	-	0.00007470A		PPB	11 Oct 2018 11:34:38	S	Std
0.2 - 1	SO.2	816	-	0.20007470A		PPB	11 Oct 2018 11:36:47	S	Std
2.5 - 1	SO.5	9035	-	2.50007470A		PPB	11 Oct 2018 11:38:56	S	Std
5.0 - 1	SO.5	18192	-	5.00007470A		PPB	11 Oct 2018 11:41:04	S	Std
7.5 - 1	SO.5	27460	-	7.50007470A		PPB	11 Oct 2018 11:43:12	S	Std
10.0 - 1	SO.5	36206	-	10.00007470A		PPB	11 Oct 2018 11:52:26	S	Std
ICV17 - 1	ICV17	14327	3.9345	-7470A		PPB	11 Oct 2018 11:54:55	U	SMPL
ICB17 - 1	ICB17	-174	-0.0623	-7470A		PPB	11 Oct 2018 11:57:03	U	SMPL
CCV54 - 1	CCV54	19069	5.2415	-7470A		PPB	11 Oct 2018 11:59:12	U	SMPL
CCB54 - 1	CCB54	-257	-0.0852	-7470A		PPB	11 Oct 2018 12:01:18	U	SMPL
CRA - 1	CRA	712	0.1819	-7470A		PPB	11 Oct 2018 12:03:27	U	SMPL
HighStd - 1	HighStd	36000	9.9080	-7470A		PPB	11 Oct 2018 12:05:34	U	SMPL
ChkStd - 1	ChkStd	25558	7.0300	-7470A		PPB	11 Oct 2018 12:07:41	U	SMPL
PB113720BL - 1	PBW	-164	-0.0595	-7470A		PPB	11 Oct 2018 12:09:50	U	SMPL
PB113720BS - 1	LCSW	13087	3.5927	-7470A		PPB	11 Oct 2018 12:12:00	U	SMPL
PB113720TB - 1	PB113683TB	-143	-0.0537	-7470A		PPB	11 Oct 2018 12:14:07	U	SMPL
J5326-04 - 1	FL-PIT-1	44	-0.0022	-7470A		PPB	11 Oct 2018 12:16:16	U	SMPL
J5326-04DUP - 1	FL-PIT-1DUP	32	-0.0055	-7470A		PPB	11 Oct 2018 12:18:25	U	SMPL
J5326-04MS - 1	FL-PIT-1MS	12421	3.4092	-7470A		PPB	11 Oct 2018 12:20:33	U	SMPL
J5326-04MSD - 1	FL-PIT-1MSD	12222	3.3543	-7470A		PPB	11 Oct 2018 12:22:40	U	SMPL
CCV55 - 1	CCV55	18031	4.9554	-7470A		PPB	11 Oct 2018 12:24:46	U	SMPL
CCB55 - 1	CCB55	-224	-0.0761	-7470A		PPB	11 Oct 2018 12:26:55	U	SMPL
J5326-08 - 1	FL-PIT-2	86	0.0094	-7470A		PPB	11 Oct 2018 12:29:02	U	SMPL
J5326-12 - 1	FL-PIT-3	19	-0.0091	-7470A		PPB	11 Oct 2018 12:31:09	U	SMPL
J5326-16 - 1	FL-PIT-4	193	0.0389	-7470A		PPB	11 Oct 2018 12:33:16	U	SMPL
J5326-20 - 1	FL-PIT-5	203	0.0416	-7470A		PPB	11 Oct 2018 12:35:23	U	SMPL
J5326-24 - 1	FL-PIT-6	194	0.0391	-7470A		PPB	11 Oct 2018 12:37:30	U	SMPL
J5326-28 - 1	FL-PIT-7	-2	-0.0149	-7470A		PPB	11 Oct 2018 12:39:38	U	SMPL
J5326-32 - 1	FL-PIT-8	77	0.0069	-7470A		PPB	11 Oct 2018 12:41:46	U	SMPL
J5326-35 - 1	FL-PH-2-STONE	52	0.0000	-7470A		PPB	11 Oct 2018 12:43:55	U	SMPL
J5326-38 - 1	FL-PH-1-STONE	23	-0.0080	-7470A		PPB	11 Oct 2018 12:46:03	U	SMPL
J5326-41 - 1	FL-PH-3-STONE	27	-0.0069	-7470A		PPB	11 Oct 2018 12:48:10	U	SMPL
CCV56 - 1	CCV56	17952	4.9336	-7470A		PPB	11 Oct 2018 12:50:17	U	SMPL
CCB56 - 1	CCB56	-138	-0.0524	-7470A		PPB	11 Oct 2018 12:52:23	U	SMPL
J5366-03 - 1	RBR200035	9	-0.0118	-7470A		PPB	11 Oct 2018 12:54:30	U	SMPL
J5369-10 - 1	CONCRETE-PILE	44	-0.0022	-7470A		PPB	11 Oct 2018 12:56:37	U	SMPL
J5372-02 - 1	OILY-DEBRI	81	0.0080	-7470A		PPB	11 Oct 2018 12:58:44	U	SMPL
J5373-02 - 1	RB37708-7211990-45911RT	73	0.0058	-7470A		PPB	11 Oct 2018 13:00:52	U	SMPL
J5373-04 - 1	RT1499-3783-RT3026	120	0.0187	-7470A		PPB	11 Oct 2018 13:02:59	U	SMPL
PB113721BL - 1	PBW	-18	-0.0193	-7470A		PPB	11 Oct 2018 13:05:06	U	SMPL
PB113721BS - 1	LCSW	12130	3.3289	-7470A		PPB	11 Oct 2018 13:07:14	U	SMPL
PB113721TB - 1	PB113689TB	-93	-0.0400	-7470A		PPB	11 Oct 2018 13:09:23	U	SMPL
J5365-02 - 1	CY-1	12	-0.0110	-7470A		PPB	11 Oct 2018 13:11:32	U	SMPL
J5365-02DUP - 1	CY-1DUP	20	-0.0088	-7470A		PPB	11 Oct 2018 13:13:39	U	SMPL
CCV57 - 1	CCV57	17993	4.9449	-7470A		PPB	11 Oct 2018 13:15:47	U	SMPL
CCB57 - 1	CCB57	-179	-0.0637	-7470A		PPB	11 Oct 2018 13:17:54	U	SMPL
J5365-02MS - 1	CY-1MS	12570	3.4502	-7470A		PPB	11 Oct 2018 13:20:03	U	SMPL
J5365-02MSD - 1	CY-1MSD	12899	3.5409	-7470A		PPB	11 Oct 2018 13:22:10	U	SMPL
J5367-02 - 1	DS-VB-16202	-7	-0.0163	-7470A		PPB	11 Oct 2018 13:24:20	U	SMPL
PB113722BL - 1	PBW	14	-0.0105	-7470A		PPB	11 Oct 2018 13:26:27	U	SMPL
PB113722BS - 1	LCSW	12601	3.4588	-7470A		PPB	11 Oct 2018 13:28:35	U	SMPL
J5365-01 - 1	CY-1	-179	-0.0637	-7470A		PPB	11 Oct 2018 13:30:42	U	SMPL
J5367-01 - 1	DS-VB-16202	128	0.0210	-7470A		PPB	11 Oct 2018 13:32:52	U	SMPL
J5367-01DUP - 1	DS-VB-16202DUP	121	0.0190	-7470A		PPB	11 Oct 2018 13:35:01	U	SMPL
J5367-01MS - 1	DS-VB-16202MS	14919	4.0977	-7470A		PPB	11 Oct 2018 13:37:11	U	SMPL
J5367-01MSD - 1	DS-VB-16202MSD	14720	4.0428	-7470A		PPB	11 Oct 2018 13:39:18	U	SMPL
CCV58 - 1	CCV58	18098	4.9739	-7470A		PPB	11 Oct 2018 13:41:26	U	SMPL
CCB58 - 1	CCB58	-177	-0.0631	-7470A		PPB	11 Oct 2018 13:43:34	U	SMPL
J5326-04LX5 - 1		-3	-0.0152	-7470A		PPB	11 Oct 2018 13:45:42	U	SMPL
J5326-04A - 1		14158	3.8879	-7470A		PPB	11 Oct 2018 13:47:50	U	SMPL
J5365-02LX5 - 1		-210	-0.0722	-7470A		PPB	11 Oct 2018 13:49:57	U	SMPL
J5365-02A - 1		13898	3.8162	-7470A		PPB	11 Oct 2018 13:52:05	U	SMPL
J5367-01LX5 - 1		-119	-0.0471	-7470A		PPB	11 Oct 2018 13:54:12	U	SMPL
J5367-01A - 1		16373	4.4984	-7470A		PPB	11 Oct 2018 13:56:21	U	SMPL
CCV59 - 1	CCV59	18133	4.9835	-7470A		PPB	11 Oct 2018 13:58:29	U	SMPL
CCB59 - 1	CCB59	-169	-0.0609	-7470A		PPB	11 Oct 2018 14:00:39	U	SMPL