



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98545

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	10/11/2018	09:35
2	J5330-01		1	8	>100.0	>212.0	>212.0	10/11/2018	09:55
3	J5360-01		1	9	>100.0	>212.0	>212.0	10/11/2018	10:25
4	J5365-02		1	8	>100.0	>212.0	>212.0	10/11/2018	10:58
5	J5365-02DUP		1	8	>100.0	>212.0	>212.0	10/11/2018	11:15
6	J5367-02		1	7	>100.0	>212.0	>212.0	10/11/2018	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp101118

WorkList ID : 117895

Date : 10/11/2018 8:59:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/19/2018	Water	J5330-01	Flash Point	Cool 4 deg C	CHAZ02	N21	EP1-WWW01	10/05/2018	1010A
10/22/2018	Water	J5360-01	Flash Point	Cool 4 deg C	DIST01	P11	2-AVE-22	10/09/2018	1010A
10/12/2018	Water	J5365-02	Flash Point	Cool 4 deg C	PSEG01	P21	CY-1	10/09/2018	1010A
10/12/2018	Water	J5367-02	Flash Point	Cool 4 deg C	PSEG01	P21	DS-VB-16202	10/09/2018	1010A

Date/Time

Received by:

Relinquished by:

10/11/18 9:00 AM
KS
CP

Date/Time

Received by:

Relinquished by:

10/11/18 11:45 AM
CP
KS