



Analytical Summary Report

Analysis Method: SM2310 B

Reviewed By: AMANDEEP

Parameter: Acidity

Supervisor Review By: apatel

Run Number: LB98550

Constant: 50,000

Normality1 (T1): 0.02

Normality2 (T2): 0.02

pH Meter ID: WC PH METER-1

Reagent/Standard	Lot/Log #
Acidity LCSW -50 ppm	WP68650

Seq	LabID	ClientID	TV (mg/L)	DL	Sample Vol (mL)	Residual Chlorine	Initial pH	T1 (mL)	Adj pH	T2 (mL)	End Point pH	Result (mg/L)	Anal Date	Anal Time
1	LB98550BL	LB98550BL		1	50.00	NEGATIVE	5.29	5.00	2.96	5.02	8.20	0.40	10/11/2018	12:57
2	LB98550BS	LB98550BS	-50	1	50.00	NEGATIVE	9.98	5.00	3.20	2.98	8.21	-40.40	10/11/2018	13:10
3	J4880-29	PT-ACD-WP		1	50.00	NEGATIVE	4.06	5.00	3.86	82.40	8.24	1,548.0	10/11/2018	13:36

T1 = Titrant1 (H2SO4)

T2 = Titrant2 (NaOH)

Result = ((T2 * Normality2) - (T1 * Normality1)) * Constant / Sample Vol