


ISM02.4

LB 98553
INSTRUMENT ID: CV2

Linear

 μ Abs.:

41531



A= 0.0000e+000

B= 2.4273e-004

C= -2.7005e-004

Rho= 0.9998929

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.019	0.019	78	0.000	78					-
0.2	0.200	0.210	0.010	868	0.0 %	868					5
2.5	2.500	2.526	0.026	10409	0.0 %	10409					i
5.0	5.000	4.949	-0.051	20388	0.0 %	20388					-1
7.5	7.500	7.415	-0.085	30550	0.0 %	30550					-1
10.0	10.000	10.081	0.081	41531	0.0 %	41531					1

AB
10/11/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		78	-	0	ISM02.4	PPB	10/11/2018 14:15	Std
0.2		868	-	0.2	ISM02.4	PPB	10/11/2018 14:17	Std
2.5		10409	-	2.5	ISM02.4	PPB	10/11/2018 14:20	Std
5		20388	-	5	ISM02.4	PPB	10/11/2018 14:22	Std
7.5		30550	-	7.5	ISM02.4	PPB	10/11/2018 14:24	Std
10		41531	-	10	ISM02.4	PPB	10/11/2018 14:27	Std
ICV	ICV	15936	3.8679	-	ISM02.4	PPB	10/11/2018 14:30	SMPL
ICB	ICB	-119	-0.0292	-	ISM02.4	PPB	10/11/2018 14:32	SMPL
CCV060	CCV060	20808	5.0505	-	ISM02.4	PPB	10/11/2018 14:34	SMPL
CCB060	CCB060	-167	-0.0408	-	ISM02.4	PPB	10/11/2018 14:36	SMPL
PB113731BL	PBW003	-57	-0.0141	-	ISM02.4	PPB	10/11/2018 14:39	SMPL
J5288-01	MH0093	12	0.0026	-	ISM02.4	PPB	10/11/2018 14:41	SMPL
J5288-02	MH0094	-89	-0.0219	-	ISM02.4	PPB	10/11/2018 14:43	SMPL
J5288-03	MH0095	-29	-0.0073	-	ISM02.4	PPB	10/11/2018 14:46	SMPL
J5288-04	MH0096	44	0.0104	-	ISM02.4	PPB	10/11/2018 14:48	SMPL
J5288-05	MH0097	-48	-0.0119	-	ISM02.4	PPB	10/11/2018 14:50	SMPL
J5288-06	MH0122	16	0.0036	-	ISM02.4	PPB	10/11/2018 14:52	SMPL
J5288-07	MH0123	-20	-0.0051	-	ISM02.4	PPB	10/11/2018 14:55	SMPL
J5288-08	MH0098	-61	-0.0151	-	ISM02.4	PPB	10/11/2018 14:57	SMPL
J5288-09	MH0099	-42	-0.0105	-	ISM02.4	PPB	10/11/2018 14:59	SMPL
J5288-10	MH0100	-10	-0.0027	-	ISM02.4	PPB	10/11/2018 15:02	SMPL
J5288-11	MH0101	-51	-0.0126	-	ISM02.4	PPB	10/11/2018 15:04	SMPL
J5288-12	MH0102	-17	-0.0044	-	ISM02.4	PPB	10/11/2018 15:06	SMPL
J5288-13	MH0103	-21	-0.0054	-	ISM02.4	PPB	10/11/2018 15:08	SMPL
J5288-14	MH0116	29	0.0068	-	ISM02.4	PPB	10/11/2018 15:11	SMPL
J5288-15	MH0117	140	0.0337	-	ISM02.4	PPB	10/11/2018 15:13	SMPL
J5288-16	MH0118	-30	-0.0076	-	ISM02.4	PPB	10/11/2018 15:15	SMPL
J5288-17	MH0119	-17	-0.0044	-	ISM02.4	PPB	10/11/2018 15:17	SMPL
J5288-18	MH0120	-19	-0.0049	-	ISM02.4	PPB	10/11/2018 15:20	SMPL
J5288-19	MH0111	-14	-0.0037	-	ISM02.4	PPB	10/11/2018 15:22	SMPL
CCV061	CCV061	20948	5.0845	-	ISM02.4	PPB	10/11/2018 15:24	SMPL
CCB061	CCB061	-131	-0.0321	-	ISM02.4	PPB	10/11/2018 15:27	SMPL
J5288-20	MH0111D	-34	-0.0085	-	ISM02.4	PPB	10/11/2018 15:29	SMPL
J5288-21	MH0111S	4862	1.1799	-	ISM02.4	PPB	10/11/2018 15:31	SMPL
J5288-22	MH0113	30	0.007	-	ISM02.4	PPB	10/11/2018 15:33	SMPL
PB113732BL	PBW003	75	0.0179	-	ISM02.4	PPB	10/11/2018 15:36	SMPL
J5289-01	MH5093	10	0.0022	-	ISM02.4	PPB	10/11/2018 15:38	SMPL
J5289-02	MH5094	11	0.0024	-	ISM02.4	PPB	10/11/2018 15:40	SMPL
J5289-03	MH5095	13	0.0029	-	ISM02.4	PPB	10/11/2018 15:42	SMPL
J5289-04	MH5096	-7	-0.002	-	ISM02.4	PPB	10/11/2018 15:45	SMPL
J5289-05	MH5097	63	0.015	-	ISM02.4	PPB	10/11/2018 15:47	SMPL
J5289-06	MH5122	-18	-0.0046	-	ISM02.4	PPB	10/11/2018 15:49	SMPL
J5289-07	MH5123	-30	-0.0076	-	ISM02.4	PPB	10/11/2018 15:52	SMPL
J5289-08	MH5098	71	0.017	-	ISM02.4	PPB	10/11/2018 15:54	SMPL
J5289-09	MH5099	-10	-0.0027	-	ISM02.4	PPB	10/11/2018 15:56	SMPL
J5289-10	MH5100	1	0	-	ISM02.4	PPB	10/11/2018 15:58	SMPL

INSTRUMENT ID : CV1

J5289-11	MH5101	-25	-0.0063 -	ISM02.4	PPB	10/11/2018 16:01	SMPL
J5289-12	MH5102	44	0.0104 -	ISM02.4	PPB	10/11/2018 16:03	SMPL
J5289-13	MH5103	113	0.0272 -	ISM02.4	PPB	10/11/2018 16:05	SMPL
J5289-14	MH5116	9	0.0019 -	ISM02.4	PPB	10/11/2018 16:08	SMPL
J5289-15	MH5117	0	-0.0003 -	ISM02.4	PPB	10/11/2018 16:10	SMPL
J5289-16	MH5118	-4	-0.0012 -	ISM02.4	PPB	10/11/2018 16:12	SMPL
CCV062	CCV062	20787	5.0454 -	ISM02.4	PPB	10/11/2018 16:14	SMPL
CCB062	CCB062	-145	-0.0355 -	ISM02.4	PPB	10/11/2018 16:17	SMPL
J5289-17	MH5119	33	0.0077 -	ISM02.4	PPB	10/11/2018 16:19	SMPL
J5289-18	MH5120	-43	-0.0107 -	ISM02.4	PPB	10/11/2018 16:21	SMPL
J5289-19	MH5111	13	0.0029 -	ISM02.4	PPB	10/11/2018 16:24	SMPL
J5289-20	MH5111D	0	-0.0003 -	ISM02.4	PPB	10/11/2018 16:26	SMPL
J5289-21	MH5111S	4002	0.9712 -	ISM02.4	PPB	10/11/2018 16:28	SMPL
J5289-22	MH5113	-17	-0.0044 -	ISM02.4	PPB	10/11/2018 16:30	SMPL
PB113733BL	PBW003	116	0.0279 -	ISM02.4	PPB	10/11/2018 16:33	SMPL
J5226-01	MA3YT3	2	0.0002 -	ISM02.4	PPB	10/11/2018 16:35	SMPL
J5290-01	MH0104	79	0.0189 -	ISM02.4	PPB	10/11/2018 16:37	SMPL
J5290-02	MH0105	-2	-0.0008 -	ISM02.4	PPB	10/11/2018 16:40	SMPL
J5290-03	MH0106	182	0.0439 -	ISM02.4	PPB	10/11/2018 16:42	SMPL
J5290-04	MH0107	-33	-0.0083 -	ISM02.4	PPB	10/11/2018 16:44	SMPL
J5290-05	MH0108	-11	-0.0029 -	ISM02.4	PPB	10/11/2018 16:46	SMPL
J5290-06	MH0109	-7	-0.002 -	ISM02.4	PPB	10/11/2018 16:49	SMPL
J5290-07	MH0112	-23	-0.0059 -	ISM02.4	PPB	10/11/2018 16:51	SMPL
J5290-08	MH0114	0	-0.0003 -	ISM02.4	PPB	10/11/2018 16:53	SMPL
J5290-09	MH0114D	77	0.0184 -	ISM02.4	PPB	10/11/2018 16:56	SMPL
J5290-10	MH0114S	4091	0.9928 -	ISM02.4	PPB	10/11/2018 16:58	SMPL
J5290-11	MH0115	28	0.0065 -	ISM02.4	PPB	10/11/2018 17:00	SMPL
J5297-09	MA3Z49	4	0.0007 -	ISM02.4	PPB	10/11/2018 17:02	SMPL
CCV063	CCV063	21029	5.1042 -	ISM02.4	PPB	10/11/2018 17:05	SMPL
CCB063	CCB063	-125	-0.0306 -	ISM02.4	PPB	10/11/2018 17:07	SMPL
PB113734BL	PBW003	172	0.0415 -	ISM02.4	PPB	10/11/2018 17:09	SMPL
J5292-01	MH5104	-1	-0.0005 -	ISM02.4	PPB	10/11/2018 17:12	SMPL
J5292-02	MH5105	1	0 -	ISM02.4	PPB	10/11/2018 17:14	SMPL
J5292-03	MH5106	26	0.006 -	ISM02.4	PPB	10/11/2018 17:16	SMPL
J5292-04	MH5107	-75	-0.0185 -	ISM02.4	PPB	10/11/2018 17:18	SMPL
J5292-05	MH5108	-46	-0.0114 -	ISM02.4	PPB	10/11/2018 17:21	SMPL
J5292-06	MH5109	-26	-0.0066 -	ISM02.4	PPB	10/11/2018 17:23	SMPL
J5292-07	MH5112	71	0.017 -	ISM02.4	PPB	10/11/2018 17:25	SMPL
J5292-08	MH5114	27	0.0063 -	ISM02.4	PPB	10/11/2018 17:28	SMPL
J5292-09	MH5114D	-27	-0.0068 -	ISM02.4	PPB	10/11/2018 17:30	SMPL
J5292-10	MH5114S	4135	1.0034 -	ISM02.4	PPB	10/11/2018 17:32	SMPL
J5292-11	MH5115	-56	-0.0139 -	ISM02.4	PPB	10/11/2018 17:34	SMPL
CCV064	CCV064	21147	5.1328 -	ISM02.4	PPB	10/11/2018 17:37	SMPL
CCB064	CCB064	-129	-0.0316 -	ISM02.4	PPB	10/11/2018 17:39	SMPL