



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB98558

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	10/11/2018	14:10
2	CAL2	1	Water			22.3	7.00	10/11/2018	14:12
3	CAL3	1	Water			22.1	10.04	10/11/2018	14:14
4	ICV	1	Water			22.2	7.01	10/11/2018	14:16
5	CCV1	1	Water			22.4	2.00	10/11/2018	14:18
6	J5401-01	1	Water			21.6	5.25	10/11/2018	14:20
7	J5401-01DUP	1	Water			21.7	5.26	10/11/2018	14:22
8	J5452-01	1	Water			24.1	4.79	10/11/2018	14:24
9	J5452-02	1	Water			23.9	4.67	10/11/2018	14:26
10	CCV2	1	Water			22.4	12.00	10/11/2018	14:28

WORKLIST(Hardcopy Internal Chain)

LB98558

WorkList Name : pH101118

WorkList ID : 117910

Date : 10/11/2018 11:12:31 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/04/2018	Water	J5401-01	pH	Cool 4 deg C	STER15	P63	GILROY-OMEGA-CELL	10/04/2018	SM 4500-H B
10/08/2018	Water	J5452-01	pH	Cool 4 deg C	STER16	P43	MAIN-CELL	10/08/2018	SM 4500-H B
10/08/2018	Water	J5452-02	pH	Cool 4 deg C	STER16	P43	EX-CELL	10/08/2018	SM 4500-H B

Date/Time 10/11/18 2:00
 Received by: SP
 Relinquished by: CP

Date/Time 10/11/18 2:30
 Received by: CP
 Relinquished by: SP