



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB98559

Slope : 97.7

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	10/11/2018	16:20
2	CAL2	1	Water			22.7	7.00	10/11/2018	16:22
3	CAL3	1	Water			22.5	10.04	10/11/2018	16:24
4	ICV	1	Water			22.6	7.01	10/11/2018	16:26
5	CCV1	1	Water			22.7	2.00	10/11/2018	16:28
6	J5449-01	1	Solid	20.02	20	22.7	6.94	10/11/2018	16:30
7	J5449-01DUP	1	Solid	20.03	20	22.8	6.95	10/11/2018	16:32
8	CCV2	1	Water			22.7	12.00	10/11/2018	16:34

WORKLIST(Hardcopy Internal Chain)

LB98559

WorkList Name : pH101118-1

WorkList ID : 117935

Date : 10/11/2018 3:26:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2018	Solid	J5449-01	pH	Cool 4 deg C	CTIA01	P23	WJ-18-10-09-TSHG-16R	10/11/2018	9045D

Date/Time 10/11/18 4:10
 Received by: SP
 Relinquished by: CP

Date/Time 10/11/18 4:40
 Received by: CP
 Relinquished by: SP