

LB98567


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

35125



A= 0.0000e+000

B= 2.8630e-004

C= -1.3944e-002

Rho= 0.9998706

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	27	0.000	27				
0.05	0.050	0.061	0.011	261	0.0 %	261				
0.2	0.200	0.182	-0.018	685	0.0 %	685				
2.5	2.500	2.500	0.000	8781	0.0 %	8781				
5.0	5.000	5.091	0.091	17830	0.0 %	17830				
7.5	7.500	7.380	-0.120	25825	0.0 %	25825				
10.0	10.000	10.042	0.042	35125	0.0 %	35125				

LB98567 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Oct 2018 18:11:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.0000	245.1	PPB	11 Oct 2018 18:12:15	S	Std
0.05 - 1	50.05	261	-	0.0500	245.1	PPB	11 Oct 2018 18:14:33	S	Std
0.2 - 1	50.2	685	-	0.2000	245.1	PPB	11 Oct 2018 18:16:50	S	Std
2.5 - 1	52.5	8781	-	2.5000	245.1	PPB	11 Oct 2018 18:19:08	S	Std
5.0 - 1	55	17830	-	5.0000	245.1	PPB	11 Oct 2018 18:21:25	S	Std
7.5 - 1	57.5	25825	-	7.5000	245.1	PPB	11 Oct 2018 18:23:42	S	Std
10.0 - 1	510	35125	-	10.0000	245.1	PPB	11 Oct 2018 18:28:38	S	Std
ICV18 - 1	ICV18	13637	3.8904	-245.1		PPB	11 Oct 2018 18:31:48	U	SMPL
ICB18 - 1	ICB18	-206	-0.0729	-245.1		PPB	11 Oct 2018 18:34:04	U	SMPL
CCV65 - 1	CCV65	17841	5.0940	-245.1		PPB	11 Oct 2018 18:36:21	U	SMPL
CCB65 - 1	CCB65	-172	-0.0632	-245.1		PPB	11 Oct 2018 18:38:37	U	SMPL
CRA - 1	CRA	191	0.0407	-245.1		PPB	11 Oct 2018 18:40:53	U	SMPL
HighStd - 1	HighStd	35417	10.1261	-245.1		PPB	11 Oct 2018 18:43:09	U	SMPL
ChkStd - 1	ChkStd	24611	7.0323	-245.1		PPB	11 Oct 2018 18:45:25	U	SMPL
PB113724BL - 1	PBW	11	-0.0108	-245.1		PPB	11 Oct 2018 18:50:20	U	SMPL
PB113724BS - 1	LCSW	13964	3.9840	-245.1		PPB	11 Oct 2018 18:52:36	U	SMPL
J5330-01 - 1	EP1-WW01	2071	0.5790	-245.1		PPB	11 Oct 2018 18:54:53	U	SMPL
J5330-01DUP - 1	EP1-WW01DUP	1838	0.5123	-245.1		PPB	11 Oct 2018 18:57:12	U	SMPL
J5330-01MS - 1	EP1-WW01MS	16676	4.7605	-245.1		PPB	11 Oct 2018 18:59:32	U	SMPL
J5330-01MSD - 1	EP1-WW01MSD	17047	4.8667	-245.1		PPB	11 Oct 2018 19:01:49	U	SMPL
J5360-01 - 1	2-AVE-22	-245	-0.0841	-245.1		PPB	11 Oct 2018 19:04:06	U	SMPL
CCV66 - 1	CCV66	17698	5.0531	-245.1		PPB	11 Oct 2018 19:06:24	U	SMPL
CCB66 - 1	CCB66	-167	-0.0618	-245.1		PPB	11 Oct 2018 19:08:39	U	SMPL
J5371-01 - 1	100818-JETT-F-D-B	-44	-0.0265	-245.1		PPB	11 Oct 2018 19:10:56	U	SMPL
J5371-02 - 1	100818-JETT-F-D-M	-101	-0.0429	-245.1		PPB	11 Oct 2018 19:13:13	U	SMPL
J5371-03 - 1	100818-JETT-F-D-S	-38	-0.0248	-245.1		PPB	11 Oct 2018 19:15:29	U	SMPL
J5371-04 - 1	100818-JETT-F-U-B	-41	-0.0257	-245.1		PPB	11 Oct 2018 19:17:45	U	SMPL
J5371-05 - 1	100818-JETT-F-U-M	28	-0.0059	-245.1		PPB	11 Oct 2018 19:20:02	U	SMPL
J5371-06 - 1	100818-JETT-F-U-S	20	-0.0082	-245.1		PPB	11 Oct 2018 19:22:19	U	SMPL
J5371-07 - 1	100818-TIPC-F-D-B	-40	-0.0254	-245.1		PPB	11 Oct 2018 19:24:37	U	SMPL
J5371-08 - 1	100818-TIPC-F-D-S	24	-0.0071	-245.1		PPB	11 Oct 2018 19:26:54	U	SMPL
J5371-09 - 1	100818-TIPC-F-U-B	36	-0.0036	-245.1		PPB	11 Oct 2018 19:29:12	U	SMPL
J5371-10 - 1	100818-TIPC-F-U-S	-44	-0.0265	-245.1		PPB	11 Oct 2018 19:31:28	U	SMPL
CCV67 - 1	CCV67	17597	5.0241	-245.1		PPB	11 Oct 2018 19:33:45	U	SMPL
CCB67 - 1	CCB67	-153	-0.0577	-245.1		PPB	11 Oct 2018 19:36:01	U	SMPL
J5371-11 - 1	100818-SIDI-E-D-B	-67	-0.0331	-245.1		PPB	11 Oct 2018 19:38:19	U	SMPL
J5371-12 - 1	100818-SIDI-E-D-M	-15	-0.0182	-245.1		PPB	11 Oct 2018 19:40:35	U	SMPL
J5371-13 - 1	100818-SIDI-E-D-S	-44	-0.0265	-245.1		PPB	11 Oct 2018 19:42:51	U	SMPL
J5371-14 - 1	100818-SIDI-E-U-B	25	-0.0068	-245.1		PPB	11 Oct 2018 19:45:08	U	SMPL
J5371-15 - 1	100818-SIDI-E-U-M	67	0.0052	-245.1		PPB	11 Oct 2018 19:47:24	U	SMPL
J5371-16 - 1	100818-SIDI-E-U-S	-9	-0.0165	-245.1		PPB	11 Oct 2018 19:49:42	U	SMPL
J5330-01LX5 - 1		189	0.0402	-245.1		PPB	11 Oct 2018 19:51:59	U	SMPL
J5330-01A - 1		17400	4.9677	-245.1		PPB	11 Oct 2018 19:54:16	U	SMPL
CCV68 - 1	CCV68	17666	5.0439	-245.1		PPB	11 Oct 2018 19:56:34	U	SMPL
CCB68 - 1	CCB68	-114	-0.0466	-245.1		PPB	11 Oct 2018 19:58:52	U	SMPL