

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/12/2018

TEMP1 IN: 103 10/11/2018 14:00 TEMP OUT: 104 10/11/2018 15:10

TEMP2 IN: 104 10/12/2018 16:00 TEMP OUT: 104 10/12/2018 07:00

TEMP3 IN: 105 10/12/2018 08:30 TEMP OUT: 105 10/12/2018 10:00

TEMP4 IN: 105 10/12/2018 11:00 TEMP OUT: 105 10/12/2018 12:40

Run Number: LB98568

BalanceID: wc-sc-3

OvenID: wc-oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98568BL	LB98568BL	1.4087	1.4087	100	1.4087	1.4087	1.4087	0.0000	0
2	LB98568BS	LB98568BS	1.3946	1.3946	100	1.4498	1.4498	1.4498	0.0552	552
3	J5434-01	SPDES-01D	1.3950	1.3950	900	1.3953	1.3953	1.3953	0.0003	0.3
4	J5434-02	SPDES-006	1.4001	1.4001	900	1.4026	1.4026	1.4026	0.0025	2.8
5	J5444-01	101018-SIDI-F-D-B	1.4158	1.4158	200	1.4245	1.4245	1.4245	0.0087	43.5
6	J5444-01DUP	101018-SIDI-F-D-BDUP	1.4135	1.4135	200	1.4223	1.4223	1.4223	0.0088	44
7	J5444-02	101018-SIDI-F-D-M	1.4122	1.4122	300	1.4222	1.4222	1.4222	0.0100	33.3
8	J5444-03	101018-SIDI-F-D-S	1.4049	1.4049	300	1.4140	1.4140	1.4140	0.0091	30.3
9	J5444-04	101018-SIDI-F-U-B	1.4002	1.4002	300	1.4136	1.4136	1.4136	0.0134	44.7
10	J5444-05	101018-SIDI-F-U-M	1.4016	1.4016	300	1.4121	1.4121	1.4121	0.0105	35
11	J5444-06	101018-SIDI-F-U-S	1.3954	1.3954	300	1.4030	1.4030	1.4030	0.0076	25.3
12	J5444-07	101018-DBRI-F-D-B	1.3968	1.3968	300	1.4103	1.4103	1.4103	0.0135	45
13	J5444-08	101018-DBRI-F-D-S	1.3906	1.3906	300	1.4033	1.4033	1.4033	0.0127	42.3
14	J5444-09	101018-DBRI-F-U-B	1.3987	1.3987	300	1.4144	1.4144	1.4144	0.0157	52.3
15	J5444-10	101018-DBRI-F-U-S	1.4000	1.4000	300	1.4088	1.4088	1.4088	0.0088	29.3



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 10/12/2018TEMP1 IN: 103 10/11/2018 14:00 TEMP OUT: 104 10/11/2018 15:10TEMP2 IN: 104 10/12/2018 16:00 TEMP OUT: 104 10/12/2018 07:00TEMP3 IN: 105 10/12/2018 08:30 TEMP OUT: 105 10/12/2018 10:00TEMP4 IN: 105 10/12/2018 11:00 TEMP OUT: 105 10/12/2018 12:40Run Number: LB98568BalanceID: wc-sc-3OvenID: wc-oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5444-11	101018-TIPC-F-D-B	1.4096	1.4096	300	1.4298	1.4298	1.4298	0.0202	67.3
17	J5444-12	101018-TIPC-F-D-S	1.3967	1.3967	200	1.4109	1.4109	1.4109	0.0142	71
18	J5444-13	101018-TIPC-F-U-B	1.3992	1.3992	200	1.4138	1.4138	1.4138	0.0146	73
19	J5444-14	101018-TIPC-F-U-S	1.4014	1.4014	200	1.4143	1.4143	1.4143	0.0129	64.5
20	J5444-14DUP	101018-TIPC-F-U-SDUP	1.4117	1.4117	200	1.4245	1.4245	1.4245	0.0128	64
21	J5446-01	SPDES-1	1.4140	1.4140	900	1.4149	1.4149	1.4149	0.0009	1
22	J5455-05	MS-OILY-WATER	1.4133	1.4133	300	1.4252	1.4252	1.4252	0.0119	39.7
23	J5456-01	MS-FRAC-TANK	1.4047	1.4047	10	2.1914	2.1914	2.1914	0.7867	78670

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS101218

WorkList ID : 117953

Date : 10/12/2018 7:50:44 AM

213 98568

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/17/2018	Water	J5434-01 A	TSS	Cool 4 deg C	FOST03	P23	SPDES-01D	10/10/2018	SM2540 D
10/17/2018	Water	J5434-02 A	TSS	Cool 4 deg C	FOST03	P23	SPDES-006	10/10/2018	SM2540 D
10/17/2018	Water	J5444-01 C	TSS	Cool 4 deg C	HDR103	P52	101018-SIDI-F-D-B	10/10/2018	SM2540 D
10/17/2018	Water	J5444-02 C	TSS	Cool 4 deg C	HDR103	P52	101018-SIDI-F-D-M	10/10/2018	SM2540 D
10/17/2018	Water	J5444-03 C	TSS	Cool 4 deg C	HDR103	P52	101018-SIDI-F-D-S	10/10/2018	SM2540 D
10/17/2018	Water	J5444-04 C	TSS	Cool 4 deg C	HDR103	P52	101018-SIDI-F-U-B	10/10/2018	SM2540 D
10/17/2018	Water	J5444-05 C	TSS	Cool 4 deg C	HDR103	P52	101018-SIDI-F-U-M	10/10/2018	SM2540 D
10/17/2018	Water	J5444-06 C	TSS	Cool 4 deg C	HDR103	P52	101018-SIDI-F-U-S	10/10/2018	SM2540 D
10/17/2018	Water	J5444-07 C	TSS	Cool 4 deg C	HDR103	P52	101018-DBRI-F-D-B	10/10/2018	SM2540 D
10/17/2018	Water	J5444-08 C	TSS	Cool 4 deg C	HDR103	P52	101018-DBRI-F-D-S	10/10/2018	SM2540 D
10/17/2018	Water	J5444-09 C	TSS	Cool 4 deg C	HDR103	P52	101018-DBRI-F-U-B	10/10/2018	SM2540 D
10/17/2018	Water	J5444-10 C	TSS	Cool 4 deg C	HDR103	P52	101018-DBRI-F-U-S	10/10/2018	SM2540 D
10/15/2018	Water	J5444-11 C	TSS	Cool 4 deg C	HDR103	P52	101018-TIPC-F-D-B	10/10/2018	SM2540 D
10/15/2018	Water	J5444-12 C	TSS	Cool 4 deg C	HDR103	P52	101018-TIPC-F-D-S	10/10/2018	SM2540 D
10/15/2018	Water	J5444-13 C	TSS	Cool 4 deg C	HDR103	P52	101018-TIPC-F-U-B	10/10/2018	SM2540 D
10/15/2018	Water	J5444-14 C	TSS	Cool 4 deg C	HDR103	P52	101018-TIPC-F-U-S	10/10/2018	SM2540 D
10/12/2018	Water	J5446-01 P	TSS	Cool 4 deg C	SKAN06	P31	SPDES-1	10/11/2018	SM2540 D
10/16/2018	Water	J5455-05 Q	TSS	Cool 4 deg C	PSEG01	P23	MS-OILY-WATER	10/11/2018	SM2540 D
10/16/2018	Water	J5456-01 F	TSS	Cool 4 deg C	PSEG01	P23	MS-FRAC-TANK	10/11/2018	SM2540 D

FP-9823888

Date/Time 10/12/18 08:00
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 10/12/18 12:30
 Received by: [Signature]
 Relinquished by: [Signature]