



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB98569

Slope : 97.7

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	10/11/2018	16:20
2	CAL2	1	Water			22.7	7.00	10/11/2018	16:22
3	CAL3	1	Water			22.5	10.04	10/11/2018	16:24
4	ICV	1	Water			22.6	7.01	10/11/2018	16:26
5	CCV1	1	Water			22.6	2.00	10/11/2018	17:08
6	J5393-12	1	Water			22.4	9.42	10/11/2018	17:10
7	J5455-06	1	Water			23.8	5.92	10/11/2018	17:14
8	J5456-02	1	Water			22.9	7.65	10/11/2018	17:16
9	J5456-02DUP	1	Water			23.0	7.66	10/11/2018	17:18
10	CCV2	1	Water			22.6	12.00	10/11/2018	17:20

WORKLIST(Hardcopy Internal Chain)

LB 98569

WorkList Name : corrosivity101118-1

WorkList ID : 117939

Date : 10/11/2018 4:58:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/10/2018	Water	J5393-12	Corrosivity	Cool 4 deg C	TETR06	P31	WE13-GW-TANK3-IDW-10102	10/10/2018	9040C
10/11/2018	Water	J5455-06	Corrosivity	Cool 4 deg C	PSEG01	P23	MS-OILY-WATER	10/11/2018	9040C
10/11/2018	Water	J5456-02	Corrosivity	Cool 4 deg C	PSEG01	P23	MS-FRAC-TANK	10/11/2018	9040C

Date/Time 10/11/18 5:00
 Received by: SP
 Relinquished by: CP

Date/Time 10/11/18 5:25
 Received by: CP
 Relinquished by: SP