



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB98571

Slope : 97.7

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	10/11/2018	16:20
2	CAL2	1	Water			22.7	7.00	10/11/2018	16:22
3	CAL3	1	Water			22.5	10.04	10/11/2018	16:24
4	ICV	1	Water			22.6	7.01	10/11/2018	16:26
5	CCV1	1	Water			22.5	2.00	10/11/2018	17:40
6	J5193-02	1	Solid	20.02	20	24.3	7.38	10/11/2018	17:42
7	J5193-02DUP	1	Solid	20.03	20	24.4	7.39	10/11/2018	17:44
8	J5197-02	1	Solid	20.02	20	23.9	7.02	10/11/2018	17:46
9	J5199-04	1	Solid	20.04	20	24.1	7.65	10/11/2018	17:48
10	J5205-04	1	Solid	20.04	20	24.2	6.42	10/11/2018	17:50
11	J5395-02	1	Solid	20.02	20	24.1	6.72	10/11/2018	17:52
12	J5450-03	1	Solid	20.02	20	24.1	7.28	10/11/2018	17:54
13	J5457-03	1	Solid	20.03	20	24.0	5.84	10/11/2018	17:56
14	CCV2	1	Water			22.6	12.00	10/11/2018	17:58

WORKLIST(Hardcopy Internal Chain)

LB98571

WorkList Name : corrosivity101118

WorkList ID : 117938

Date : 10/11/2018 4:58:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2018	Solid	J5193-02	Corrosivity	Cool 4 deg C	PSEG05	P23	BU-01-101118	10/11/2018	9045D
10/11/2018	Solid	J5197-02	Corrosivity	Cool 4 deg C	PSEG05	P21	HD-01-101118	10/11/2018	9045D
10/11/2018	Solid	J5199-04	Corrosivity	Cool 4 deg C	PSEG05		OK-02-101118	10/11/2018	9045D
10/11/2018	Solid	J5205-04	Corrosivity	Cool 4 deg C	PSEG05		TR-04-101118	10/11/2018	9045D
10/11/2018	Solid	J5395-02	Corrosivity	Cool 4 deg C	PSEG01	P22	PH-SOIL	10/11/2018	9045D
10/11/2018	Solid	J5450-03	Corrosivity	Cool 4 deg C	PSEG01	P21	VNJ218-210-282	10/11/2018	9045D
10/11/2018	Solid	J5457-03	Corrosivity	Cool 4 deg C	PSEG01	P23	KG-SOIL	10/11/2018	9045D

Date/Time 10/11/18 5:35
 Received by: SP
 Relinquished by: CP

Date/Time 10/11/18 6:05
 Received by: GP
 Relinquished by: SP