



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98576

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	81.2	10/12/2018	09:35
2	J5455-06		1	6	>100.0	>212.0	>212.0	10/12/2018	10:02
3	J5455-06DUP		1	6	>100.0	>212.0	>212.0	10/12/2018	10:40
4	J5456-02		1	7	>100.0	>212.0	>212.0	10/12/2018	11:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB98576

WorkList Name : fp101218

WorkList ID : 117966

Date : 10/12/2018 9:06:33 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/16/2018	Water	J5455-06 B	Flash Point	Cool 4 deg C	PSEG01	P23	MS-OILY-WATER	10/11/2018	1010A
10/16/2018	Water	J5456-02 B	Flash Point	Cool 4 deg C	PSEG01	P23	MS-FRAC-TANK	10/11/2018	1010A

Date/Time 10/12/18 9:10 AM
 Received by: KS
 Relinquished by: CP

Date/Time 10/12/18 11:10 AM
 Received by: CP
 Relinquished by: KS