



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB98581

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.01	10/04/2018	16:20
2	CAL2	1	Water			22.6	7.00	10/04/2018	16:22
3	CAL3	1	Water			22.7	10.04	10/04/2018	16:24
4	ICV	1	Water			22.4	7.01	10/04/2018	16:26
5	CCV1	1	Water			22.6	2.00	10/04/2018	16:28
6	J5255-03	1	Solid	20.02	20	22.8	8.62	10/04/2018	16:30
7	J5255-03DUP	1	Solid	20.03	20	22.9	8.63	10/04/2018	16:32
8	J5255-06	1	Solid	20.02	20	23.1	9.05	10/04/2018	16:34
9	CCV2	1	Water			22.8	12.00	10/04/2018	16:36

WORKLIST(Hardcopy Internal Chain)

LB98581

WorkList Name : CORROSIONITY100418-2

WorkList ID : 117974

Date : 10/04/2018 4:05:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/03/2018	Solid	J5255-03	Corrosivity	Cool 4 deg C	TECT01		TRACK-D-1	10/03/2018	9045D
10/03/2018	Solid	J5255-06	Corrosivity	Cool 4 deg C	TECT01		TRACK-D-2	10/03/2018	9045D

Date/Time 10/04/18 4:10
 Received by: SP
 Relinquished by: CP

Date/Time 10/04/18 4:40
 Received by: CP
 Relinquished by: SP