



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB98582

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	10/05/2018	16:40
2	CAL2	1	Water			22.4	7.00	10/05/2018	16:42
3	CAL3	1	Water			22.8	10.04	10/05/2018	16:44
4	ICV	1	Water			22.6	7.01	10/05/2018	16:46
5	CCV1	1	Water			22.7	2.00	10/05/2018	16:48
6	J5287-03	1	Solid	20.02	20	22.9	9.08	10/05/2018	16:50
7	J5287-03DUP	1	Solid	20.03	20	23.0	9.09	10/05/2018	16:52
8	J5287-06	1	Solid	20.02	20	23.5	8.45	10/05/2018	16:54
9	J5287-09	1	Solid	20.04	20	22.8	7.26	10/05/2018	16:56
10	J5287-12	1	Solid	20.02	20	22.5	5.86	10/05/2018	16:58
11	J5287-15	1	Solid	20.02	20	23.6	6.29	10/05/2018	17:00
12	J5287-18	1	Solid	20.03	20	23.1	8.66	10/05/2018	17:02
13	CCV2	1	Water			22.8	12.00	10/05/2018	17:04

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY100518-2

WorkList ID : 117975

Date : 10/05/2018 4:10:15 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/04/2018	Solid	J5287-03	Corrosivity	Cool 4 deg C	TECT01		TRACK-D-3	10/04/2018	9045D
10/04/2018	Solid	J5287-06	Corrosivity	Cool 4 deg C	TECT01		TRACK-D-4	10/04/2018	9045D
10/04/2018	Solid	J5287-09	Corrosivity	Cool 4 deg C	TECT01		TRACK-D-5	10/04/2018	9045D
10/04/2018	Solid	J5287-12	Corrosivity	Cool 4 deg C	TECT01		TRACK-D-6	10/04/2018	9045D
10/04/2018	Solid	J5287-15	Corrosivity	Cool 4 deg C	TECT01		TRACK-D-7	10/04/2018	9045D
10/04/2018	Solid	J5287-18	Corrosivity	Cool 4 deg C	TECT01		TRACK-D-8	10/04/2018	9045D

Date/Time 10/05/18 4:30

Received by: SP

Relinquished by: CP

Date/Time 10/05/18 5:10

Received by: CP

Relinquished by: SP