

Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB98583

Reagent/Standard	Lot/Loq #
MBAS CAL STD, 0.5PPM	WP68684
MBAS CCV STD, 1.0PPM	WP68683
MBAS ICV-LCS std 1.0PPM	WP68693
MBAS DAILY CHECK STD, 0.01PPM	WP68699
MBAS DAILY CHECK STD, 2.0PPM	WP68700
MBAS CAL STD 2 PPM	WP68685
MBAS CAL STD 1 PPM	WP68687
MBAS CAL STD 0.1 PPM	WP68690
mbas cal std 0.05 ppm	WP68691
MBAS CAL STD 0.01 PPM	WP68692
MBAS BLANK 0 PPM	WP68682
MBAS CAL STD 0.4 PPM	WP68688
MBAS CAL STD 0.2 PPM	WP68689
Methylene blue reagent-1000ml	WP66778
MBAS Wash Solution	WP66800

Intercept: 0.0007

Slope: 0.2219

Regression: 0.9999

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Anal Date	Anal Time
1	CAL1	0	1	0.000	10/12/2018	12:04
2	CAL2	0.01	1	0.003	10/12/2018	12:05
3	CAL3	0.05	1	0.011	10/12/2018	12:05
4	CAL4	0.1	1	0.023	10/12/2018	12:06
5	CAL5	0.2	1	0.043	10/12/2018	12:06
6	CAL6	0.4	1	0.088	10/12/2018	12:07
7	CAL7	0.5	1	0.116	10/12/2018	12:07
8	CAL8	1	1	0.224	10/12/2018	12:08
9	CAL9	2	1	0.441	10/12/2018	12:08



Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB98583

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Result (mg/L)	AnalDate	Anal Time
1	ICV	1	1	0.209	0.939	10/12/2018	12:09
2	ICB		1	0.000	-0.003	10/12/2018	12:09
3	CHK1	0.01	1	0.003	0.010	10/12/2018	12:10
4	CHK2	2	1	0.407	1.831	10/12/2018	12:11
5	CCV1	1	1	0.216	0.970	10/12/2018	12:12
6	CCB1		1	0.000	-0.003	10/12/2018	12:12
7	LB98583BL		1	0.000	-0.003	10/12/2018	12:13
8	LB98583BS	1	1	0.211	0.948	10/12/2018	12:13
9	J4880-30		1	0.164	0.736	10/12/2018	12:14
10	J4880-30DUP		1	0.165	0.740	10/12/2018	12:14
11	CCV2	1	1	0.218	0.979	10/12/2018	12:15
12	CCB2		1	0.000	-0.003	10/12/2018	12:15