

245.1


L398596

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38402



A= 0.0000e+000

B= 2.6479e-004

C= 1.1853e-002

Rho= 0.9994416

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.020	0.020	32	0.000	32				
0.05	0.050	0.057	0.007	169	0.0 %	169				
0.2	0.200	0.213	0.013	758	0.0 %	758				
2.5	2.500	2.620	0.120	9848	0.0 %	9848				
5.0	5.000	4.867	-0.133	18336	0.0 %	18336				
7.5	7.500	7.294	-0.206	27500	0.0 %	27500				
10.0	10.000	10.180	0.180	38402	0.0 %	38402				

LB98596 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Oct 2018 17:32:47

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	32	-	0.0000	245.1	PPB	12 Oct 2018 17:34:08	S	Std
0.05 - 1	50.05	169	-	0.0500	245.1	PPB	12 Oct 2018 17:36:26	S	Std
0.2 - 1	50.2	758	-	0.2000	245.1	PPB	12 Oct 2018 17:38:44	S	Std
2.5 - 1	52.5	9848	-	2.5000	245.1	PPB	12 Oct 2018 17:48:35	S	Std
5.0 - 1	55	18336	-	5.0000	245.1	PPB	12 Oct 2018 17:50:52	S	Std
7.5 - 1	57.5	27500	-	7.5000	245.1	PPB	12 Oct 2018 17:55:51	S	Std
10.0 - 1	510	38402	-	10.0000	245.1	PPB	12 Oct 2018 18:00:43	S	Std
ICV21 - 1	ICV21	14314	3.8021	-245.1		PPB	12 Oct 2018 18:06:05	U	SMPL
ICB21 - 1	ICB21	-169	-0.0329	-245.1		PPB	12 Oct 2018 18:08:21	U	SMPL
CCV76 - 1	CCV76	17452	4.6330	-245.1		PPB	12 Oct 2018 18:10:38	U	SMPL
CCB76 - 1	CCB76	-201	-0.0414	-245.1		PPB	12 Oct 2018 18:12:54	U	SMPL
CRA - 1	CRA	137	0.0481	-245.1		PPB	12 Oct 2018 18:15:11	U	SMPL
HighStd - 1	HighStd	37729	10.0021	-245.1		PPB	12 Oct 2018 18:17:27	U	SMPL
ChkStd - 1	ChkStd	26015	6.9004	-245.1		PPB	12 Oct 2018 18:19:43	U	SMPL
PB113815BL - 1	PBW	-302	-0.0681	-245.1		PPB	12 Oct 2018 18:22:02	U	SMPL
PB113815BS - 1	LCSW	13391	3.5577	-245.1		PPB	12 Oct 2018 18:24:21	U	SMPL
J5383-01 - 1	100918-SIDI-F-D-B	-154	-0.0289	-245.1		PPB	12 Oct 2018 18:26:38	U	SMPL
J5383-02 - 1	100918-SIDI-F-D-M	-42	0.0007	-245.1		PPB	12 Oct 2018 18:28:55	U	SMPL
J5383-03 - 1	100918-SIDI-F-D-S	-43	0.0005	-245.1		PPB	12 Oct 2018 18:31:13	U	SMPL
J5383-04 - 1	100918-SIDI-F-U-B	-12	0.0087	-245.1		PPB	12 Oct 2018 18:33:31	U	SMPL
J5383-05 - 1	100918-SIDI-F-U-M	-34	0.0029	-245.1		PPB	12 Oct 2018 18:35:48	U	SMPL
CCV77 - 1	CCV77	17657	4.6873	-245.1		PPB	12 Oct 2018 18:38:03	U	SMPL
CCB77 - 1	CCB77	-194	-0.0395	-245.1		PPB	12 Oct 2018 18:40:19	U	SMPL
J5383-06 - 1	100918-SIDI-F-U-S	-61	-0.0043	-245.1		PPB	12 Oct 2018 18:42:37	U	SMPL
J5383-07 - 1	100918-TIPC-F-D-B	-12	0.0087	-245.1		PPB	12 Oct 2018 18:44:53	U	SMPL
J5383-08 - 1	100918-TIPC-F-D-S	-12	0.0087	-245.1		PPB	12 Oct 2018 18:47:09	U	SMPL
J5383-09 - 1	100918-TIPC-F-U-B	14	0.0156	-245.1		PPB	12 Oct 2018 18:49:25	U	SMPL
J5383-10 - 1	100918-TIPC-F-U-S	24	0.0182	-245.1		PPB	12 Oct 2018 18:51:42	U	SMPL
J5383-11 - 1	100918-DBRI-F-D-B	-17	0.0074	-245.1		PPB	12 Oct 2018 18:53:59	U	SMPL
J5383-12 - 1	100918-DBRI-F-D-S	-22	0.0060	-245.1		PPB	12 Oct 2018 18:56:17	U	SMPL
J5383-13 - 1	100918-DBRI-F-U-B	-26	0.0050	-245.1		PPB	12 Oct 2018 18:58:34	U	SMPL
J5383-14 - 1	100918-DBRI-F-U-S	54	0.0262	-245.1		PPB	12 Oct 2018 19:00:52	U	SMPL
J5383-14DUP - 1	100918-DBRI-F-U-SDUP	-8	0.0097	-245.1		PPB	12 Oct 2018 19:03:08	U	SMPL
CCV78 - 1	CCV78	17733	4.7074	-245.1		PPB	12 Oct 2018 19:05:25	U	SMPL
CCB78 - 1	CCB78	-187	-0.0377	-245.1		PPB	12 Oct 2018 19:07:41	U	SMPL
J5383-14MS - 1	100918-DBRI-F-U-SMS	14800	3.9308	-245.1		PPB	12 Oct 2018 19:09:59	U	SMPL
J5383-14MSD - 1	100918-DBRI-F-U-SMSD	14451	3.8383	-245.1		PPB	12 Oct 2018 19:12:16	U	SMPL
PB113816BL - 1	PBW	-124	-0.0210	-245.1		PPB	12 Oct 2018 19:14:32	U	SMPL
PB113816BS - 1	LCSW	14007	3.7208	-245.1		PPB	12 Oct 2018 19:16:49	U	SMPL
J5444-01 - 1	101018-SIDI-F-D-B	-163	-0.0313	-245.1		PPB	12 Oct 2018 19:19:05	U	SMPL
J5444-02 - 1	101018-SIDI-F-D-M	-49	-0.0011	-245.1		PPB	12 Oct 2018 19:21:22	U	SMPL
J5444-03 - 1	101018-SIDI-F-D-S	-24	0.0055	-245.1		PPB	12 Oct 2018 19:23:39	U	SMPL
J5444-04 - 1	101018-SIDI-F-U-B	-44	0.0002	-245.1		PPB	12 Oct 2018 19:25:57	U	SMPL
J5444-05 - 1	101018-SIDI-F-U-M	-10	0.0092	-245.1		PPB	12 Oct 2018 19:28:16	U	SMPL
J5444-06 - 1	101018-SIDI-F-U-S	-39	0.0015	-245.1		PPB	12 Oct 2018 19:30:32	U	SMPL
CCV79 - 1	CCV79	17974	4.7712	-245.1		PPB	12 Oct 2018 19:32:49	U	SMPL
CCB79 - 1	CCB79	-220	-0.0464	-245.1		PPB	12 Oct 2018 19:35:05	U	SMPL
J5444-07 - 1	101018-DBRI-F-D-B	-50	-0.0014	-245.1		PPB	12 Oct 2018 19:37:23	U	SMPL
J5444-08 - 1	101018-DBRI-F-D-S	-53	-0.0022	-245.1		PPB	12 Oct 2018 19:39:40	U	SMPL
J5444-09 - 1	101018-DBRI-F-U-B	-42	0.0007	-245.1		PPB	12 Oct 2018 19:41:56	U	SMPL
J5444-10 - 1	101018-DBRI-F-U-S	-64	-0.0051	-245.1		PPB	12 Oct 2018 19:44:12	U	SMPL
J5444-11 - 1	101018-TIPC-F-D-B	-24	0.0055	-245.1		PPB	12 Oct 2018 19:46:29	U	SMPL
J5444-12 - 1	101018-TIPC-F-D-S	-27	0.0047	-245.1		PPB	12 Oct 2018 19:48:46	U	SMPL
J5444-13 - 1	101018-TIPC-F-U-B	-24	0.0055	-245.1		PPB	12 Oct 2018 19:51:04	U	SMPL
J5444-14 - 1	101018-TIPC-F-U-S	-31	0.0036	-245.1		PPB	12 Oct 2018 19:53:21	U	SMPL
J5444-14DUP - 1	101018-TIPC-F-U-SDUP	-51	-0.0017	-245.1		PPB	12 Oct 2018 19:55:39	U	SMPL
J5444-14MS - 1	101018-TIPC-F-U-SMS	12855	3.4157	-245.1		PPB	12 Oct 2018 19:57:56	U	SMPL
CCV80 - 1	CCV80	17522	4.6515	-245.1		PPB	12 Oct 2018 20:00:14	U	SMPL
CCB80 - 1	CCB80	-184	-0.0369	-245.1		PPB	12 Oct 2018 20:02:30	U	SMPL
J5444-14MSD - 1	101018-TIPC-F-U-SMSD	12793	3.3993	-245.1		PPB	12 Oct 2018 20:04:48	U	SMPL
J5383-14LX5 - 1		-167	-0.0324	-245.1		PPB	12 Oct 2018 20:07:04	U	SMPL
J5383-14A - 1		16436	4.3640	-245.1		PPB	12 Oct 2018 20:09:21	U	SMPL
J5444-14LX5 - 1		-181	-0.0361	-245.1		PPB	12 Oct 2018 20:11:38	U	SMPL
J5444-14A - 1		15917	4.2265	-245.1		PPB	12 Oct 2018 20:13:54	U	SMPL
CCV81 - 1	CCV81	18644	4.9486	-245.1		PPB	12 Oct 2018 20:16:11	U	SMPL
CCB81 - 1	CCB81	-218	-0.0459	-245.1		PPB	12 Oct 2018 20:18:29	U	SMPL