

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/15/2018

TEMP1 IN: 103 10/15/2018 07:00 TEMP OUT: 104 10/15/2018 08:05

TEMP2 IN: 104 10/15/2018 09:00 TEMP OUT: 104 10/15/2018 10:05

TEMP3 IN: 105 10/15/2018 11:00 TEMP OUT: 105 10/15/2018 12:40

TEMP4 IN: 105 10/15/2018 13:15 TEMP OUT: 105 10/15/2018 15:00

Run Number: LB98599

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98599BL	LB98599BL	1.3951	1.3951	100	1.3951	1.3951	1.3951	0.0000	0
2	LB98599BS	LB98599BS	1.3909	1.3909	100	1.4410	1.4410	1.4410	0.0501	501
3	J5458-01	RAV-RAINEY-BEFORE-10-03-18	1.3979	1.3979	400	1.3999	1.3999	1.3999	0.0020	5
4	J5458-01DUP	RAV-RAINEY-BEFORE-10-03-18D	1.3941	1.3941	400	1.3960	1.3960	1.3960	0.0019	4.8
5	J5458-02	RAV-RAINEY-AFTER-10-03-18	1.3878	1.3878	500	1.3915	1.3915	1.3915	0.0037	7.4
6	J5458-03	RAV-006-10-03-18	1.3950	1.3950	900	1.3977	1.3977	1.3977	0.0027	3
7	J5458-04	RAV-01D-10-03-18	1.3897	1.3897	900	1.3902	1.3902	1.3902	0.0005	0.6
8	J5458-05	RAV-01A-10-03-18	1.3928	1.3928	900	1.3973	1.3973	1.3973	0.0045	5
9	J5458-06	RAV-011-10-03-18	1.4101	1.4101	500	1.4156	1.4156	1.4156	0.0055	11
10	J5511-01	101118-SIDI-F-D-B	1.3878	1.3878	200	1.4014	1.4014	1.4014	0.0136	68
11	J5511-02	101118-SIDI-F-D-M	1.3871	1.3871	300	1.4009	1.4009	1.4009	0.0138	46
12	J5511-03	101118-SIDI-F-D-S	1.3926	1.3926	400	1.4056	1.4056	1.4056	0.0130	32.5
13	J5511-04	101118-SIDI-F-U-B	1.4029	1.4029	200	1.4234	1.4234	1.4234	0.0205	102.5
14	J5511-05	101118-SIDI-F-U-M	1.3882	1.3882	200	1.4037	1.4037	1.4037	0.0155	77.5
15	J5511-06	101118-SIDI-F-U-S	1.4058	1.4058	400	1.4189	1.4189	1.4189	0.0131	32.8



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/15/2018

TEMP1 IN: 103 10/15/2018 07:00 TEMP OUT: 104 10/15/2018 08:05

TEMP2 IN: 104 10/15/2018 09:00 TEMP OUT: 104 10/15/2018 10:05

TEMP3 IN: 105 10/15/2018 11:00 TEMP OUT: 105 10/15/2018 12:40

TEMP4 IN: 105 10/15/2018 13:15 TEMP OUT: 105 10/15/2018 15:00

Run Number: LB98599

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5511-07	101118-TIPC-F-D-B	1.3936	1.3936	300	1.4117	1.4117	1.4117	0.0181	60.3
17	J5511-08	101118-TIPC-F-D-S	1.3894	1.3894	200	1.4006	1.4006	1.4006	0.0112	56
18	J5511-09	101118-TIPC-F-U-B	1.3929	1.3929	200	1.4082	1.4082	1.4082	0.0153	76.5
19	J5511-10	101118-TIPC-F-U-S	1.4013	1.4013	200	1.4123	1.4123	1.4123	0.0110	55
20	J5511-10DUP	101118-TIPC-F-U-SDUP	1.3926	1.3926	200	1.4035	1.4035	1.4035	0.0109	54.5

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss101518

WorkList ID : 118008

Date : 10/15/2018 7:57:10 AM

11808599

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/10/2018	Water	J5458-01	B TSS	Cool 4 deg C	FOST03	P41	RAV-RAINEY-BEFORE-10-03	10/03/2018	SM2540 D
10/10/2018	Water	J5458-02	B TSS	Cool 4 deg C	FOST03	P41	RAV-RAINEY-AFTER-10-03-1	10/03/2018	SM2540 D
10/10/2018	Water	J5458-03	B TSS	Cool 4 deg C	FOST03	P41	RAV-006-10-03-18	10/03/2018	SM2540 D
10/10/2018	Water	J5458-04	A TSS	Cool 4 deg C	FOST03	P41	RAV-01D-10-03-18	10/03/2018	SM2540 D
10/10/2018	Water	J5458-05	A TSS	Cool 4 deg C	FOST03	P41	RAV-01A-10-03-18	10/03/2018	SM2540 D
10/10/2018	Water	J5458-06	A TSS	Cool 4 deg C	FOST03	P41	RAV-011-10-03-18	10/03/2018	SM2540 D
10/18/2018	Water	J5511-01	C TSS	Cool 4 deg C	HDR103	D21	101118-SIDI-F-D-B	10/11/2018	SM2540 D
10/18/2018	Water	J5511-02	C TSS	Cool 4 deg C	HDR103	D21	101118-SIDI-F-D-M	10/11/2018	SM2540 D
10/18/2018	Water	J5511-03	C TSS	Cool 4 deg C	HDR103	D21	101118-SIDI-F-D-S	10/11/2018	SM2540 D
10/18/2018	Water	J5511-04	C TSS	Cool 4 deg C	HDR103	D21	101118-SIDI-F-U-B	10/11/2018	SM2540 D
10/18/2018	Water	J5511-05	C TSS	Cool 4 deg C	HDR103	D21	101118-SIDI-F-U-M	10/11/2018	SM2540 D
10/18/2018	Water	J5511-06	C TSS	Cool 4 deg C	HDR103	D21	101118-SIDI-F-U-S	10/11/2018	SM2540 D
10/16/2018	Water	J5511-07	C TSS	Cool 4 deg C	HDR103	D21	101118-TIPC-F-D-B	10/11/2018	SM2540 D
10/16/2018	Water	J5511-08	C TSS	Cool 4 deg C	HDR103	D21	101118-TIPC-F-D-S	10/11/2018	SM2540 D
10/16/2018	Water	J5511-09	C TSS	Cool 4 deg C	HDR103	D21	101118-TIPC-F-U-B	10/11/2018	SM2540 D
10/16/2018	Water	J5511-10	C TSS	Cool 4 deg C	HDR103	D21	101118-TIPC-F-U-S	10/11/2018	SM2540 D

FP-9823888

Date/Time 10/15/18, 08:00
 Received by: AS
 Relinquished by: AS

Date/Time 10/15/18, 12:00
 Received by: CS
 Relinquished by: AS