



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB98601

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.5	4.01	10/12/2018	16:40
2	CAL2	1	Water			22.6	7.00	10/12/2018	16:42
3	CAL3	1	Water			22.6	10.04	10/12/2018	16:44
4	ICV	1	Water			22.3	7.01	10/12/2018	16:46
5	CCV1	1	Water			22.5	2.00	10/12/2018	16:48
6	J5198-12	1	Solid	20.02	20	22.3	8.36	10/12/2018	16:50
7	J5198-12DUP	1	Solid	20.03	20	22.4	8.37	10/12/2018	16:52
8	J5198-14	1	Solid	20.02	20	22.2	8.48	10/12/2018	16:54
9	J5198-16	1	Solid	20.04	20	22.2	7.01	10/12/2018	16:56
10	J5198-18	1	Solid	20.02	20	21.9	7.69	10/12/2018	16:58
11	J5198-20	1	Solid	20.02	20	21.3	8.27	10/12/2018	17:00
12	J5509-03	1	Solid	20.03	20	22.3	5.12	10/12/2018	17:02
13	CCV2	1	Water			22.6	12.00	10/12/2018	17:04

WORKLIST(Hardcopy Internal Chain)

LB98601

WorkList Name : corrosivity101218

WorkList ID : 117994

Date : 10/12/2018 4:16:42 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/12/2018	Solid	J5198-12	Corrosivity	Cool 4 deg C	PSEG05		CL-02-101218-A	10/12/2018	9045D
10/12/2018	Solid	J5198-14	Corrosivity	Cool 4 deg C	PSEG05		CL-02-101218-B	10/12/2018	9045D
10/12/2018	Solid	J5198-16	Corrosivity	Cool 4 deg C	PSEG05		CL-02-101218-C	10/12/2018	9045D
10/12/2018	Solid	J5198-18	Corrosivity	Cool 4 deg C	PSEG05		CL-02-101218-D	10/12/2018	9045D
10/12/2018	Solid	J5198-20	Corrosivity	Cool 4 deg C	PSEG05		CL-02-101218-E	10/12/2018	9045D
10/12/2018	Solid	J5509-03	Corrosivity	Cool 4 deg C	PSEG01	P31	MS-SOIL	10/12/2018	9045D

Date/Time 10/15/18 4:35
 Received by: SP
 Relinquished by: CP

Date/Time 10/12/18 5:20
 Received by: CP
 Relinquished by: SP