

LB98613

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/15/2018 13:55

Test: Total CN ug/h

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.568	0.0	0.089	
ICB1	-0.835	0.0	0.004	
CCV1	245.450	0.0	0.223	
CCB1	-0.840	0.0	0.004	
PB113843BL	-0.850	0.0	0.004	
PB113843BS	96.397	0.0	0.090	
LOWPB113843	8.714	0.0	0.012	
HIGHPB113843	490.843	0.0	0.442	
J5198-11	-0.187	0.0	0.004	
J5198-11DUP	-0.558	0.0	0.004	
J5198-11MS	36.044	0.0	0.036	
J5198-11MSD	35.762	0.0	0.036	
J5198-13	-0.485	0.0	0.004	
J5198-15	-0.521	0.0	0.004	
CCV2	245.458	0.0	0.223	
CCB2	-1.104	0.0	0.003	
J5198-17	-0.430	0.0	0.004	
J5509-02	-0.858	0.0	0.004	
J5198-19	-0.774	0.0	0.004	
CCV3	262.191	0.0	0.238	
CCB3	-1.261	0.0	0.003	

N 21
Mean 71.749
SD 130.9499
CV% 182.51

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : ds

10/15/2018 10:34


Test Total CN

Accepted 10/15/2018 10:34

Factor 1120
Bias 0.004

Coeff. of det. 0.999826

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.3242	0.0000	
2	5.0PPBCN	0.008	4.1378	5.0000	
3	10PPBCN	0.012	8.9168	10.0000	
4	50PPBCN	0.049	49.9219	50.0000	
5	100PPBCN	0.095	101.0776	100.0000	
6	250PPBCN	0.232	254.8950	250.0000	
7	500PPBCN	0.448	497.3751	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Mon Oct 15 13:25:32 2018

Mon Oct 15 13:51:49 2018



Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-1.3242	µg/l	10/15/2018 10:34:22	
5.0PPBCN	A	Total CN	P	4.1378	µg/l	10/15/2018 10:34:23	
10PPBCN	A	Total CN	P	8.9168	µg/l	10/15/2018 10:34:24	
50PPBCN	A	Total CN	P	49.9219	µg/l	10/15/2018 10:34:25	
100PPBCN	A	Total CN	P	101.0776	µg/l	10/15/2018 10:34:26	
250PPBCN	A	Total CN	P	254.895	µg/l	10/15/2018 10:34:27	
500PPBCN	A	Total CN	P	497.3751	µg/l	10/15/2018 10:34:28	
ICV1	S	Total CN	P	94.5682	µg/l	10/15/2018 13:25:32	
ICB1	S	Total CN	P	-0.8348	µg/l	10/15/2018 13:25:33	
CCV1	S	Total CN	P	245.4499	µg/l	10/15/2018 13:25:34	
CCB1	S	Total CN	P	-0.8398	µg/l	10/15/2018 13:25:35	
PB113843BL	S	Total CN	P	-0.8498	µg/l	10/15/2018 13:25:36	
PB113843BS	S	Total CN	P	96.3973	µg/l	10/15/2018 13:25:37	
LOWPB113843	S	Total CN	P	8.7138	µg/l	10/15/2018 13:25:38	
HIGHPB113843	S	Total CN	P	490.8428	µg/l	10/15/2018 13:25:39	
J5198-11	S	Total CN	P	-0.187	µg/l	10/15/2018 13:25:40	
J5198-11DUP	S	Total CN	P	-0.5584	µg/l	10/15/2018 13:25:41	
J5198-11MS	S	Total CN	P	36.0441	µg/l	10/15/2018 13:25:42	
J5198-11MSD	S	Total CN	P	35.762	µg/l	10/15/2018 13:33:07	
J5198-13	S	Total CN	P	-0.4851	µg/l	10/15/2018 13:33:08	
J5198-15	S	Total CN	P	-0.521	µg/l	10/15/2018 13:33:09	
CCV2	S	Total CN	P	245.458	µg/l	10/15/2018 13:33:10	
CCB2	S	Total CN	P	-1.1037	µg/l	10/15/2018 13:33:11	
J5198-17	S	Total CN	P	-0.4298	µg/l	10/15/2018 13:33:12	
J5509-02	S	Total CN	P	-0.8582	µg/l	10/15/2018 13:33:14	
J5198-19	S	Total CN	P	-0.7739	µg/l	10/15/2018 13:51:47	
CCV3	S	Total CN	P	262.1913	µg/l	10/15/2018 13:51:48	
CCB3	S	Total CN	P	-1.2607	µg/l	10/15/2018 13:51:49	