


245.1

LB 98620
INSTRUMENT ID: CV2

Linear

 μ Abs.:

30983



A= 0.0000e+000

B= 3.2101e-004

C= -1.8393e-002

Rho= 0.9999223

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	-12	0.000	-12				
0.05	0.050	0.047	-0.003	205	0.0 %	205				
0.2	0.200	0.191	-0.009	653	0.0 %	653				
2.5	2.500	2.478	-0.022	7776	0.0 %	7776				
5.0	5.000	5.078	0.078	15876	0.0 %	15876				
7.5	7.500	7.551	0.051	23579	0.0 %	23579				
10.0	10.000	9.927	-0.073	30983	0.0 %	30983				

LB98620 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Oct 2018 17:09:27

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-12	-	0.0000	245.1	PPB	15 Oct 2018 17:14:20	S	Std
0.05 - 1	50.05	205	-	0.0500	245.1	PPB	15 Oct 2018 17:16:37	S	Std
0.2 - 1	50.2	653	-	0.2000	245.1	PPB	15 Oct 2018 17:18:55	S	Std
2.5 - 1	52.5	7776	-	2.5000	245.1	PPB	15 Oct 2018 17:21:12	S	Std
5.0 - 1	55	15876	-	5.0000	245.1	PPB	15 Oct 2018 17:23:30	S	Std
7.5 - 1	57.5	23579	-	7.5000	245.1	PPB	15 Oct 2018 17:25:46	S	Std
10.0 - 1	510	30983	-	10.0000	245.1	PPB	15 Oct 2018 17:33:19	S	Std
ICV23 - 1	ICV23	12550	4.0102	-245.1		PPB	15 Oct 2018 17:36:04	U	SMPL
ICB23 - 1	ICB23	-17	-0.0239	-245.1		PPB	15 Oct 2018 17:40:54	U	SMPL
CCV88 - 1	CCV88	16062	5.1376	-245.1		PPB	15 Oct 2018 17:43:09	U	SMPL
CCB88 - 1	CCB88	-177	-0.0752	-245.1		PPB	15 Oct 2018 17:45:25	U	SMPL
CRA - 1	CRA	193	0.0436	-245.1		PPB	15 Oct 2018 18:02:55	U	SMPL
HighStd - 1	HighStd	30206	9.6779	-245.1		PPB	15 Oct 2018 18:05:10	U	SMPL
ChkStd - 1	ChkStd	22174	7.0996	-245.1		PPB	15 Oct 2018 18:07:27	U	SMPL
PB113853BL - 1	PBW	182	0.0400	-245.1		PPB	15 Oct 2018 18:12:29	U	SMPL
PB113853BS - 1	LCSW	12477	3.9868	-245.1		PPB	15 Oct 2018 18:14:45	U	SMPL
J5164-10 - 1	LOD-MDL-WATER-04-QT4-2018	243	0.0596	-245.1		PPB	15 Oct 2018 18:22:18	U	SMPL
J5164-11 - 1	LOQ-WATER-05-QT4-2018	1389	0.4275	-245.1		PPB	15 Oct 2018 18:27:14	U	SMPL
J5164-12 - 1	MDL-WATER-06-QT4-2018	180	0.0394	-245.1		PPB	15 Oct 2018 18:29:31	U	SMPL
J5511-01 - 1	101118-SIDI-F-D-B	469	0.1322	-245.1		PPB	15 Oct 2018 18:31:49	U	SMPL
J5511-02 - 1	101118-SIDI-F-D-M	419	0.1161	-245.1		PPB	15 Oct 2018 18:34:05	U	SMPL
CCV89 - 1	CCV89	16097	5.1488	-245.1		PPB	15 Oct 2018 18:36:21	U	SMPL
CCB89 - 1	CCB89	-157	-0.0688	-245.1		PPB	15 Oct 2018 18:38:36	U	SMPL
J5511-03 - 1	101118-SIDI-F-D-S	436	0.1216	-245.1		PPB	15 Oct 2018 18:40:53	U	SMPL
J5511-04 - 1	101118-SIDI-F-U-B	474	0.1338	-245.1		PPB	15 Oct 2018 18:43:09	U	SMPL
J5511-05 - 1	101118-SIDI-F-U-M	490	0.1389	-245.1		PPB	15 Oct 2018 18:45:25	U	SMPL
J5511-06 - 1	101118-SIDI-F-U-S	478	0.1350	-245.1		PPB	15 Oct 2018 18:47:41	U	SMPL
J5511-07 - 1	101118-TIPC-F-D-B	480	0.1357	-245.1		PPB	15 Oct 2018 18:49:58	U	SMPL
J5511-08 - 1	101118-TIPC-F-D-S	462	0.1299	-245.1		PPB	15 Oct 2018 18:52:15	U	SMPL
J5511-09 - 1	101118-TIPC-F-U-B	484	0.1370	-245.1		PPB	15 Oct 2018 18:54:32	U	SMPL
J5511-10 - 1	101118-TIPC-F-U-S	485	0.1373	-245.1		PPB	15 Oct 2018 18:56:49	U	SMPL
J5511-10DUP - 1	101118-TIPC-F-U-SDUP	502	0.1428	-245.1		PPB	15 Oct 2018 18:59:07	U	SMPL
J5511-10MS - 1	101118-TIPC-F-U-SMS	12856	4.1085	-245.1		PPB	15 Oct 2018 19:01:23	U	SMPL
CCV90 - 1	CCV90	16335	5.2252	-245.1		PPB	15 Oct 2018 19:03:40	U	SMPL
CCB90 - 1	CCB90	-193	-0.0803	-245.1		PPB	15 Oct 2018 19:05:56	U	SMPL
J5511-10MSD - 1	101118-TIPC-F-U-SMSD	13950	4.4596	-245.1		PPB	15 Oct 2018 19:08:13	U	SMPL
J5511-10LX5 - 1		0	-0.0184	-245.1		PPB	15 Oct 2018 19:10:30	U	SMPL
J5511-10A - 1		13327	4.2597	-245.1		PPB	15 Oct 2018 19:12:46	U	SMPL
CCV91 - 1	CCV91	16363	5.2342	-245.1		PPB	15 Oct 2018 19:15:03	U	SMPL
CCB91 - 1	CCB91	-50	-0.0344	-245.1		PPB	15 Oct 2018 19:17:20	U	SMPL