



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98627

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.7	4.01	10/15/2018	16:00
2	CAL2	1	Water			20.8	7.00	10/15/2018	16:02
3	CAL3	1	Water			20.6	10.04	10/15/2018	16:04
4	ICV	1	Water			20.7	7.01	10/15/2018	16:06
5	CCV1	1	Water			20.8	2.00	10/15/2018	17:10
6	J5524-04	1	Solid	20.03	20	20.5	6.78	10/15/2018	17:12
7	J5524-04DUP	1	Solid	20.04	20	20.6	6.79	10/15/2018	17:14
8	J5524-08	1	Solid	20.03	20	20.6	5.45	10/15/2018	17:16
9	J5526-11	1	Solid	20.03	20	20.8	6.46	10/15/2018	17:18
10	J5526-13	1	Solid	20.02	20	20.8	6.95	10/15/2018	17:20
11	J5526-15	1	Solid	20.02	20	20.9	6.25	10/15/2018	17:22
12	J5526-23	1	Solid	20.04	20	20.8	7.55	10/15/2018	17:24
13	J5526-25	1	Solid	20.02	20	20.7	8.08	10/15/2018	17:26
14	J5527-04	1	Solid	20.02	20	20.4	8.29	10/15/2018	17:28
15	CCV2	1	Water			20.6	12.00	10/15/2018	17:30

UB 94627

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity101518

WorkList ID : 118045

Date : 10/15/2018 3:38:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2018	Solid	J5524-04	Corrosivity	Cool 4 deg C	PSEG01	P31	BS-SOIL-1	10/15/2018	9045D
10/15/2018	Solid	J5524-08	Corrosivity	Cool 4 deg C	PSEG01	P31	BS-SOIL-2	10/15/2018	9045D
10/15/2018	Solid	J5526-11	Corrosivity	Cool 4 deg C	DIST01	D62	SUP-4B-C-101518-1	10/15/2018	9045D
10/15/2018	Solid	J5526-13	Corrosivity	Cool 4 deg C	DIST01	D62	SUP-4B-C-101518-2	10/15/2018	9045D
10/15/2018	Solid	J5526-15	Corrosivity	Cool 4 deg C	DIST01	D62	SUP-4B-C-101518-3	10/15/2018	9045D
10/15/2018	Solid	J5526-23	Corrosivity	Cool 4 deg C	DIST01	D62	SUP-4A-C-101518-1	10/15/2018	9045D
10/15/2018	Solid	J5526-25	Corrosivity	Cool 4 deg C	DIST01	D62	SUP-4A-C-101518-2	10/15/2018	9045D
10/15/2018	Solid	J5527-04	Corrosivity	Cool 4 deg C	PSEG01	P31	SP-STONE	10/15/2018	9045D

Date/Time 10/15/18 5:05
 Received by: SP
 Relinquished by: SP

Date/Time 10/15/18 5:35
 Received by: SP
 Relinquished by: SP