



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98628

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.7	4.01	10/15/2018	16:00
2	CAL2	1	Water			20.8	7.00	10/15/2018	16:02
3	CAL3	1	Water			20.6	10.04	10/15/2018	16:04
4	ICV	1	Water			20.7	7.01	10/15/2018	16:06
5	CCV1	1	Water			20.8	2.00	10/15/2018	16:08
6	J5519-02	1	Water			16.2	6.59	10/15/2018	16:10
7	J5519-02DUP	1	Water			16.3	6.60	10/15/2018	16:12
8	CCV2	1	Water			20.5	12.00	10/15/2018	16:14

LB 98628

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity101518-1

WorkList ID : 118046

Date : 10/15/2018 3:38:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2018	Water	J5519-02	Corrosivity	Cool 4 deg C	PSEG01	P23	BS-DRUM	10/15/2018	9040C

Date/Time 10/15/18 3:50
 Received by: SP
 Relinquished by: CP

Date/Time 10/15/18 4:20
 Received by: CP
 Relinquished by: SP