



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98632

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.6	10/16/2018	09:20
2	J5519-02		1	7	>100.0	>212.0	>212.0	10/16/2018	09:55
3	J5519-02DUP		1	7	>100.0	>212.0	>212.0	10/16/2018	10:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB98632

WorkList Name : fp101618

WorkList ID : 118065

Date : 10/16/2018 8:47:29 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/18/2018	Water	J5519-02	C Flash Point	Cool 4 deg C	PSEG01	P23	BS-DRUM	10/15/2018	1010A

Date/Time 10/16/18 8:50 AM
 Received by: KS
 Relinquished by: CP

Date/Time 10/16/18 10:20 AM
 Received by: CP
 Relinquished by: KS