



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98639

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	10/16/2018	15:50
2	CAL2	1	Water			22.3	7.00	10/16/2018	15:52
3	CAL3	1	Water			22.2	10.04	10/16/2018	15:54
4	ICV	1	Water			22.5	7.01	10/16/2018	15:56
5	CCV1	1	Water			22.4	2.00	10/16/2018	15:58
6	J5533-01	1	Water			18.1	5.42	10/16/2018	16:00
7	J5533-01DUP	1	Water			18.2	5.43	10/16/2018	16:02
8	J5546-01	1	Water			18.7	7.23	10/16/2018	16:04
9	J5546-03	1	Water			18.7	8.02	10/16/2018	16:06
10	CCV2	1	Water			22.4	12.00	10/16/2018	16:08

WORKLIST(Hardcopy Internal Chain)

LB 98639

WorkList Name : pHJ5533

WorkList ID : 118089

Date : 10/16/2018 2:54:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2018	Water	J5533-01 A	pH	Cool 4 deg C	STER14	D31	JS-SYSTEM	10/15/2018	SM 4500-H B
10/16/2018	Water	J5546-01 C	pH	Cool 4 deg C	TULL02	D22	BNM-1	10/16/2018	SM 4500-H B
10/16/2018	Water	J5546-03 C	pH	Cool 4 deg C	TULL02	D22	MH-20A	10/16/2018	SM 4500-H B

Date/Time 10/16/18 3:40
 Received by: SP
 Relinquished by: SP

Date/Time 10/16/18 4:15
 Received by: CP
 Relinquished by: SP