



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98643

Slope : 98.8

BalanceID: WC-SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	10/16/2018	15:50
2	CAL2	1	Water			22.3	7.00	10/16/2018	15:52
3	CAL3	1	Water			22.2	10.04	10/16/2018	15:54
4	ICV	1	Water			22.5	7.01	10/16/2018	15:56
5	CCV1	1	Water			22.4	2.00	10/16/2018	16:20
6	J5200-10	1	Solid	20.02	20	21.4	8.52	10/16/2018	16:22
7	J5200-10DUP	1	Solid	20.03	20	21.5	8.53	10/16/2018	16:24
8	J5200-12	1	Solid	20.03	20	21.2	8.94	10/16/2018	16:26
9	J5536-04	1	Solid	20.02	20	21.8	9.07	10/16/2018	16:28
10	CCV2	1	Water			22.2	12.00	10/16/2018	16:30

WORKLIST(Hardcopy Internal Chain)

LB98643

WorkList Name : corrosivity101618

WorkList ID : 118095

Date : 10/16/2018 4:00:18 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/16/2018	Solid	J5200-10	Corrosivity	Cool 4 deg C	PSEG05		OR-01-101618-A	10/16/2018	9045D
10/16/2018	Solid	J5200-12	Corrosivity	Cool 4 deg C	PSEG05		OR-01-101618-B	10/16/2018	9045D
10/16/2018	Solid	J5536-04	Corrosivity	Cool 4 deg C	PSEG01	P23	351-SOIL	10/16/2018	9045D

Date/Time 10/16/18 4:05
 Received by: SP
 Relinquished by: CP

Date/Time 10/16/18 4:35
 Received by: CP
 Relinquished by: SP