

LB 98645


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

42307



A= 0.0000e+000

B= 2.3550e-004

C= -1.6614e-002

Rho= 0.9998945

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	102	0.000	102				
0.2	0.200	0.199	-0.001	915	0.0 %	915				
2.5	2.500	2.495	-0.005	10663	0.0 %	10663				
5.0	5.000	4.946	-0.054	21071	0.0 %	21071				
7.5	7.500	7.607	0.107	32370	0.0 %	32370				
10.0	10.000	9.947	-0.053	42307	0.0 %	42307				

LB98645 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Oct 2018 17:45:51

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	102	-	0.00007471B	PPB	16 Oct 2018 17:47:21	S	Std
0.2 - 1	50.2	915	-	0.20007471B	PPB	16 Oct 2018 17:49:38	S	Std
2.5 - 1	52.5	10663	-	2.50007471B	PPB	16 Oct 2018 17:51:56	S	Std
5.0 - 1	55	21071	-	5.00007471B	PPB	16 Oct 2018 17:54:14	S	Std
7.5 - 1	57.5	32370	-	7.50007471B	PPB	16 Oct 2018 17:56:31	S	Std
10.0 - 1	510	42307	-	10.00007471B	PPB	16 Oct 2018 18:01:26	S	Std
ICV25 - 1	ICV25	16248	3.8098	-7471B	PPB	16 Oct 2018 18:06:50	U	SMPL
ICB25 - 1	ICB25	-97	-0.0395	-7471B	PPB	16 Oct 2018 18:09:05	U	SMPL
CCV07 - 1	CCV07	21188	4.9732	-7471B	PPB	16 Oct 2018 18:11:21	U	SMPL
CCB07 - 1	CCB07	-88	-0.0373	-7471B	PPB	16 Oct 2018 18:13:37	U	SMPL
CRA - 1	CRA	908	0.1972	-7471B	PPB	16 Oct 2018 18:15:54	U	SMPL
HighStd - 1	HighStd	42662	10.0304	-7471B	PPB	16 Oct 2018 18:18:09	U	SMPL
ChkStd - 1	ChkStd	29952	7.0372	-7471B	PPB	16 Oct 2018 18:20:25	U	SMPL
PB113878BL - 1	PBS	-192	-0.0618	-7471B	PPB	16 Oct 2018 18:22:44	U	SMPL
PB113878BS - 1	LCSS	15867	3.7201	-7471B	PPB	16 Oct 2018 18:25:02	U	SMPL
J5198-11 - 1	CL-02-101218-A	3922	0.9070	-7471B	PPB	16 Oct 2018 18:27:19	U	SMPL
J5198-13 - 1	CL-02-101218-B	2093	0.4763	-7471B	PPB	16 Oct 2018 18:29:36	U	SMPL
J5198-15 - 1	CL-02-101218-C	9883	2.3109	-7471B	PPB	16 Oct 2018 18:31:54	U	SMPL
J5198-17 - 1	CL-02-101218-D	4893	1.1357	-7471B	PPB	16 Oct 2018 18:34:12	U	SMPL
J5198-19 - 1	CL-02-101218-E	3079	0.7085	-7471B	PPB	16 Oct 2018 18:36:28	U	SMPL
CCV08 - 1	CCV08	20940	4.9148	-7471B	PPB	16 Oct 2018 18:38:44	U	SMPL
CCB08 - 1	CCB08	-88	-0.0373	-7471B	PPB	16 Oct 2018 18:41:00	U	SMPL
J5270-01 - 1	STOCK-PILE-A	13036	3.0534	-7471B	PPB	16 Oct 2018 18:43:16	U	SMPL
J5270-03 - 1	STOCK-PILE-B	18955	4.4474	-7471B	PPB	16 Oct 2018 18:45:32	U	SMPL
J5270-05 - 1	STOCK-PILE-C	12574	2.9446	-7471B	PPB	16 Oct 2018 18:47:47	U	SMPL
J5509-02 - 1	MS-SOIL	16816	3.9436	-7471B	PPB	16 Oct 2018 18:50:04	U	SMPL
J5516-02 - 1	SSSI-1011-S-DH-32	3032	0.6974	-7471B	PPB	16 Oct 2018 18:52:20	U	SMPL
J5516-04 - 1	SSSI-1011-S-DH-31	329	0.0609	-7471B	PPB	16 Oct 2018 18:54:37	U	SMPL
J5516-06 - 1	SSSI-1011-S-DH-30	599	0.1245	-7471B	PPB	16 Oct 2018 18:56:54	U	SMPL
J5516-08 - 1	SSSI-1012-S-DH-29	1929	0.4377	-7471B	PPB	16 Oct 2018 18:59:12	U	SMPL
J5524-01 - 1	BS-SOIL-1	1428	0.3197	-7471B	PPB	16 Oct 2018 19:01:30	U	SMPL
J5524-05 - 1	BS-SOIL-2	836	0.1803	-7471B	PPB	16 Oct 2018 19:03:47	U	SMPL
CCV09 - 1	CCV09	21263	4.9909	-7471B	PPB	16 Oct 2018 19:06:03	U	SMPL
CCB09 - 1	CCB09	-77	-0.0347	-7471B	PPB	16 Oct 2018 19:08:19	U	SMPL
J5526-10 - 1	SUP-4B-C-101518-1	1445	0.3237	-7471B	PPB	16 Oct 2018 19:10:36	U	SMPL
J5526-12 - 1	SUP-4B-C-101518-2	1885	0.4273	-7471B	PPB	16 Oct 2018 19:12:52	U	SMPL
J5526-14 - 1	SUP-4B-C-101518-3	1219	0.2705	-7471B	PPB	16 Oct 2018 19:15:09	U	SMPL
J5526-22 - 1	SUP-4A-C-101518-1	1174	0.2599	-7471B	PPB	16 Oct 2018 19:17:26	U	SMPL
J5526-24 - 1	SUP-4A-C-101518-2	2019	0.4589	-7471B	PPB	16 Oct 2018 19:19:43	U	SMPL
J5526-24DUP - 1	SUP-4A-C-101518-2DUP	2044	0.4648	-7471B	PPB	16 Oct 2018 19:22:00	U	SMPL
J5526-24MS - 1	SUP-4A-C-101518-2MS	17908	4.2008	-7471B	PPB	16 Oct 2018 19:24:18	U	SMPL
J5526-24MSD - 1	SUP-4A-C-101518-2MSD	17492	4.1028	-7471B	PPB	16 Oct 2018 19:26:35	U	SMPL
J5526-24LX5 - 1		369	0.0703	-7471B	PPB	16 Oct 2018 19:28:53	U	SMPL
J5526-24A - 1		19501	4.5759	-7471B	PPB	16 Oct 2018 19:31:09	U	SMPL
CCV10 - 1	CCV10	21443	5.0333	-7471B	PPB	16 Oct 2018 19:33:26	U	SMPL
CCB10 - 1	CCB10	-81	-0.0357	-7471B	PPB	16 Oct 2018 19:35:43	U	SMPL