


7470A

LB98649
INSTRUMENT ID: CV4

Linear

 μ Abs.:

36745



A= 0.0000e+000

B= 2.7315e-004

C= -3.6324e-003

Rho= 0.9999338

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	-19	0.000	-19				
0.2	0.200	0.218	0.018	810	0.0 %	810				
2.5	2.500	2.478	-0.022	9087	0.0 %	9087				
5.0	5.000	5.056	0.056	18522	0.0 %	18522				
7.5	7.500	7.424	-0.076	27193	0.0 %	27193				
10.0	10.000	10.033	0.033	36745	0.0 %	36745				

LB98649 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Oct 2018 09:48:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.0000	7470A	PPB	17 Oct 2018 10:02:21	S	Std
0.2 - 1	50.2	810	-	0.2000	7470A	PPB	17 Oct 2018 10:04:30	S	Std
2.5 - 1	52.5	9087	-	2.5000	7470A	PPB	17 Oct 2018 10:06:38	S	Std
5.0 - 1	55	18522	-	5.0000	7470A	PPB	17 Oct 2018 10:08:46	S	Std
7.5 - 1	57.5	27193	-	7.5000	7470A	PPB	17 Oct 2018 10:10:54	S	Std
10.0 - 1	58	36745	-	10.0000	7470A	PPB	17 Oct 2018 10:13:03	S	Std
ICV26 - 1	ICV26	14158	3.8636	-	7470A	PPB	17 Oct 2018 10:15:41	U	SMPL
ICB26 - 1	ICB26	-25	-0.0105	-	7470A	PPB	17 Oct 2018 10:20:09	U	SMPL
CCV11 - 1	CCV11	17856	4.8737	-	7470A	PPB	17 Oct 2018 10:22:15	U	SMPL
CCB11 - 1	CCB11	-146	-0.0435	-	7470A	PPB	17 Oct 2018 10:24:22	U	SMPL
CRA - 1	CRA	803	0.2157	-	7470A	PPB	17 Oct 2018 10:26:29	U	SMPL
HighStd - 1	HighStd	36286	9.9078	-	7470A	PPB	17 Oct 2018 10:28:36	U	SMPL
ChkStd - 1	ChkStd	25906	7.0725	-	7470A	PPB	17 Oct 2018 10:30:42	U	SMPL
PB113901BL - 1	PBW	-77	-0.0247	-	7470A	PPB	17 Oct 2018 10:32:51	U	SMPL
PB113901BS - 1	LCSW	14277	3.8961	-	7470A	PPB	17 Oct 2018 10:35:01	U	SMPL
PB113901TB - 1	PB113864TB	17	0.0010	-	7470A	PPB	17 Oct 2018 10:37:09	U	SMPL
PB113902BL - 1	PBW	137	0.0338	-	7470A	PPB	17 Oct 2018 10:39:18	U	SMPL
PB113902BS - 1	LCSW	14685	4.0075	-	7470A	PPB	17 Oct 2018 10:41:26	U	SMPL
PB113902TB - 1	PB113866TB	18	0.0013	-	7470A	PPB	17 Oct 2018 10:43:35	U	SMPL
PB113910BL - 1	PBW	150	0.0373	-	7470A	PPB	17 Oct 2018 10:45:42	U	SMPL
CCV12 - 1	CCV12	18053	4.9275	-	7470A	PPB	17 Oct 2018 10:47:48	U	SMPL
CCB12 - 1	CCB12	-9	-0.0061	-	7470A	PPB	17 Oct 2018 10:52:23	U	SMPL
PB113910BS - 1	LCSW	14383	3.9250	-	7470A	PPB	17 Oct 2018 10:54:29	U	SMPL
J5524-04 - 1	BS-SOIL-1	3	-0.0028	-	7470A	PPB	17 Oct 2018 10:56:37	U	SMPL
J5524-08 - 1	BS-SOIL-2	111	0.0267	-	7470A	PPB	17 Oct 2018 10:58:44	U	SMPL
J5526-11 - 1	SUP-4B-C-101518-1	114	0.0275	-	7470A	PPB	17 Oct 2018 11:00:51	U	SMPL
J5526-13 - 1	SUP-4B-C-101518-2	152	0.0379	-	7470A	PPB	17 Oct 2018 11:02:58	U	SMPL
J5526-15 - 1	SUP-4B-C-101518-3	154	0.0384	-	7470A	PPB	17 Oct 2018 11:05:06	U	SMPL
J5526-23 - 1	SUP-4A-C-101518-1	150	0.0373	-	7470A	PPB	17 Oct 2018 11:07:14	U	SMPL
J5526-25 - 1	SUP-4A-C-101518-2	153	0.0382	-	7470A	PPB	17 Oct 2018 11:09:22	U	SMPL
J5526-25DUP - 1	SUP-4A-C-101518-2DUP	127	0.0311	-	7470A	PPB	17 Oct 2018 11:11:31	U	SMPL
J5526-25MS - 1	SUP-4A-C-101518-2MS	12204	3.3298	-	7470A	PPB	17 Oct 2018 11:13:38	U	SMPL
CCV13 - 1	CCV13	17733	4.8401	-	7470A	PPB	17 Oct 2018 11:15:45	U	SMPL
CCB13 - 1	CCB13	-129	-0.0389	-	7470A	PPB	17 Oct 2018 11:17:52	U	SMPL
J5526-25MSD - 1	SUP-4A-C-101518-2MSD	12884	3.5156	-	7470A	PPB	17 Oct 2018 11:20:01	U	SMPL
J5527-04 - 1	SP-STONE	22	0.0024	-	7470A	PPB	17 Oct 2018 11:22:08	U	SMPL
J5519-02 - 1	BS-DRUM	145	0.0360	-	7470A	PPB	17 Oct 2018 11:24:15	U	SMPL
J5519-02DUP - 1	BS-DRUMDUP	194	0.0494	-	7470A	PPB	17 Oct 2018 11:28:57	U	SMPL
J5519-02MS - 1	BS-DRUMMS	12599	3.4377	-	7470A	PPB	17 Oct 2018 11:31:04	U	SMPL
J5519-02MSD - 1	BS-DRUMMSD	12453	3.3979	-	7470A	PPB	17 Oct 2018 11:33:12	U	SMPL
J5519-01 - 1	BS-DRUM	203	0.0518	-	7470A	PPB	17 Oct 2018 11:40:04	U	SMPL
J5519-01DUP - 1	BS-DRUMDUP	157	0.0393	-	7470A	PPB	17 Oct 2018 11:42:12	U	SMPL
J5519-01MS - 1	BS-DRUMMS	13167	3.5929	-	7470A	PPB	17 Oct 2018 11:44:21	U	SMPL
J5519-01MSD - 1	BS-DRUMMSD	13363	3.6464	-	7470A	PPB	17 Oct 2018 11:46:28	U	SMPL
CCV14 - 1	CCV14	17877	4.8794	-	7470A	PPB	17 Oct 2018 11:48:35	U	SMPL
CCB14 - 1	CCB14	-43	-0.0154	-	7470A	PPB	17 Oct 2018 11:53:09	U	SMPL
J5521-13 - 1	TWP-01	1451	0.3927	-	7470A	PPB	17 Oct 2018 11:55:15	U	SMPL
J5540-01 - 1	MOSAIC-FRAC-TANK-101518	392	0.1034	-	7470A	PPB	17 Oct 2018 11:57:23	U	SMPL
J5540-02 - 1	MOSAIC-MGP-SW-101518	73	0.0163	-	7470A	PPB	17 Oct 2018 11:59:30	U	SMPL
J5546-01 - 1	BNM-1	917	0.2468	-	7470A	PPB	17 Oct 2018 12:01:37	U	SMPL
J5546-03 - 1	MH-20A	126	0.0308	-	7470A	PPB	17 Oct 2018 12:03:44	U	SMPL
CCV15 - 1	CCV15	17809	4.8608	-	7470A	PPB	17 Oct 2018 12:05:52	U	SMPL
CCB15 - 1	CCB15	-114	-0.0348	-	7470A	PPB	17 Oct 2018 12:08:00	U	SMPL
J5526-25LX5 - 1		19	0.0016	-	7470A	PPB	17 Oct 2018 12:10:09	U	SMPL
J5526-25A - 1		13713	3.7420	-	7470A	PPB	17 Oct 2018 12:12:18	U	SMPL
J5519-02LX5 - 1		-63	-0.0208	-	7470A	PPB	17 Oct 2018 12:14:25	U	SMPL
J5519-02A - 1		13216	3.6063	-	7470A	PPB	17 Oct 2018 12:16:32	U	SMPL
J5519-01LX5 - 1		0	-0.0036	-	7470A	PPB	17 Oct 2018 12:18:39	U	SMPL
J5519-01A - 1		14233	3.8841	-	7470A	PPB	17 Oct 2018 12:20:46	U	SMPL
CCV16 - 1	CCV16	18051	4.9269	-	7470A	PPB	17 Oct 2018 12:22:53	U	SMPL
CCB16 - 1	CCB16	-145	-0.0432	-	7470A	PPB	17 Oct 2018 12:25:02	U	SMPL