



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB98652

Slope : 99.1

pH Meter ID : Wc pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.01	10/17/2018	12:05
2	CAL2	1	Water			22.5	7.00	10/17/2018	12:07
3	CAL3	1	Water			22.3	10.04	10/17/2018	12:09
4	ICV	1	Water			22.5	7.01	10/17/2018	12:11
5	CCV1	1	Water			22.6	2.00	10/17/2018	12:13
6	J5566-01	1	Water			21.0	5.59	10/17/2018	12:15
7	J5566-01DUP	1	Water			21.1	5.60	10/17/2018	12:17
8	CCV2	1	Water			22.4	12.00	10/17/2018	12:19

WORKLIST(Hardcopy Internal Chain)

LB98652

WorkList Name : pH101718

WorkList ID : 118128

Date : 10/17/2018 11:56:56 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/10/2018	Water	J5566-01	pH	Cool 4 deg C	STER10	D62	CELL-A-POOL-WATER(DE-IC	10/10/2018	SM 4500-H B

Date/Time 10/17/18 12:00
 Received by: SP
 Relinquished by: [Signature]

Date/Time 10/17/18 12:30
 Received by: [Signature]
 Relinquished by: SP