



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98653

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.9	10/17/2018	10:02
2	J5546-01		1	7	>100.0	>212.0	>212.0	10/17/2018	10:35
3	J5546-01DUP		1	7	>100.0	>212.0	>212.0	10/17/2018	10:55
4	J5546-03		1	8	>100.0	>212.0	>212.0	10/17/2018	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 98653

WorkList Name : FP101718

WorkList ID : 118116

Date : 10/17/2018 8:50:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/29/2018	Water	J5546-01	Flash Point	Cool 4 deg C	TULL02	D22	BNM-1	10/16/2018	1010A
10/29/2018	Water	J5546-03	Flash Point	Cool 4 deg C	TULL02	D22	MH-20A	10/16/2018	1010A

Date/Time 10/17/18 9:00 AM
 Received by: KS
 Relinquished by: RJ

Date/Time 10/17/18 11:25 AM
 Received by: CP
 Relinquished by: KS