



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98655

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.01	10/17/2018	12:05
2	CAL2	1	Water			22.5	7.00	10/17/2018	12:07
3	CAL3	1	Water			22.3	10.04	10/17/2018	12:09
4	ICV	1	Water			22.6	7.01	10/17/2018	13:46
5	CCV1	1	Water			22.4	2.00	10/17/2018	13:48
6	J5570-01	1	Solid	20.02	20	21.7	6.35	10/17/2018	13:50
7	J5570-01DUP	1	Solid	20.04	20	21.8	6.36	10/17/2018	13:52
8	J5570-02	1	Solid	20.03	20	22.0	6.53	10/17/2018	13:54
9	J5570-03	1	Solid	20.02	20	22.0	5.85	10/17/2018	13:56
10	CCV2	1	Water			22.6	12.00	10/17/2018	13:58

WORKLIST(Hardcopy Internal Chain)

LB98655

WorkList Name : pH101718-1

WorkList ID : 118131

Date : 10/17/2018 1:15:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/16/2018	Solid	J5570-01	pH	Cool 4 deg C	ENTA04	D41	TS-06	10/16/2018	9045D
10/16/2018	Solid	J5570-02	pH	Cool 4 deg C	ENTA04	D41	TS-07	10/16/2018	9045D
10/16/2018	Solid	J5570-03	pH	Cool 4 deg C	ENTA04	D41	TS-08	10/16/2018	9045D

Date/Time 10/17/18 1:27
 Received by: SP
 Relinquished by: SP

Date/Time 10/17/18 2:04
 Received by: SP
 Relinquished by: SP