


ISM02.4

LB 98660
INSTRUMENT ID: CY2

Linear

 μ Abs.:

41656



A= 0.0000e+000

B= 2.3944e-004

C= -2.8830e-002

Rho= 0.9999527

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.014	-0.014	64	0.000	64				
0.2	0.200	0.170	-0.030	831	0.0 %	831				
2.5	2.500	2.515	0.015	10626	0.0 %	10626				
5.0	5.000	5.045	0.045	21191	0.0 %	21191				
7.5	7.500	7.537	0.037	31599	0.0 %	31599				
10.0	10.000	9.945	-0.055	41656	0.0 %	41656				

900
-15
7
1
0
-1

103
10/17/18

LB98660
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	64	-	0	ISM02.4	PPB	10/17/2018 15:21	Std
	0.2 S0.2	831	-	0.2	ISM02.4	PPB	10/17/2018 15:23	Std
	2.5 S2.5	10626	-	2.5	ISM02.4	PPB	10/17/2018 15:26	Std
	5 S5	21191	-	5	ISM02.4	PPB	10/17/2018 15:28	Std
	7.5 S7.5	31599	-	7.5	ISM02.4	PPB	10/17/2018 15:30	Std
	10 S10	41656	-	10	ISM02.4	PPB	10/17/2018 15:33	Std
ICV	ICV	17296	4.1126 -		ISM02.4	PPB	10/17/2018 15:35	SMPL
ICB	ICB	-88	-0.0499 -		ISM02.4	PPB	10/17/2018 15:38	SMPL
CCV017	CCV017	21285	5.0677 -		ISM02.4	PPB	10/17/2018 15:40	SMPL
CCB017	CCB017	-82	-0.0485 -		ISM02.4	PPB	10/17/2018 15:42	SMPL
PB113903BL	PBS004	91	-0.007 -		ISM02.4	PPB	10/17/2018 15:45	SMPL
J5387-01	MYAK80	759	0.1529 -		ISM02.4	PPB	10/17/2018 15:47	SMPL
J5387-02	MYAK81	424	0.0727 -		ISM02.4	PPB	10/17/2018 15:49	SMPL
J5387-03	MYAK82	922	0.1919 -		ISM02.4	PPB	10/17/2018 15:51	SMPL
J5387-04	MYAK83	510	0.0933 -		ISM02.4	PPB	10/17/2018 15:54	SMPL
J5387-05	MYAK84	571	0.1079 -		ISM02.4	PPB	10/17/2018 15:56	SMPL
J5387-06	MYAK85	912	0.1895 -		ISM02.4	PPB	10/17/2018 15:58	SMPL
J5387-07	MYAK86	1299	0.2822 -		ISM02.4	PPB	10/17/2018 16:00	SMPL
J5387-08	MYAK87	1649	0.366 -		ISM02.4	PPB	10/17/2018 16:03	SMPL
J5387-09	MYAKK3	183	0.015 -		ISM02.4	PPB	10/17/2018 16:05	SMPL
J5387-10	MYAKK4	117	-0.0008 -		ISM02.4	PPB	10/17/2018 16:07	SMPL
J5387-11	MYAKK5	287	0.0399 -		ISM02.4	PPB	10/17/2018 16:10	SMPL
J5387-12	MYAKK6	233	0.027 -		ISM02.4	PPB	10/17/2018 16:12	SMPL
J5387-13	MYAKK7	509	0.093 -		ISM02.4	PPB	10/17/2018 16:14	SMPL
J5387-14	MYAKK8	195	0.0179 -		ISM02.4	PPB	10/17/2018 16:16	SMPL
J5387-15	MYAKK9	225	0.025 -		ISM02.4	PPB	10/17/2018 16:19	SMPL
J5387-16	MYAKL0	189	0.0164 -		ISM02.4	PPB	10/17/2018 16:21	SMPL
J5387-17	MYAKL1	237	0.0279 -		ISM02.4	PPB	10/17/2018 16:23	SMPL
J5387-18	MYAKN4	489	0.0883 -		ISM02.4	PPB	10/17/2018 16:25	SMPL
J5387-19	MYAKN5	894	0.1852 -		ISM02.4	PPB	10/17/2018 16:28	SMPL
CCV018	CCV018	21156	5.0368 -		ISM02.4	PPB	10/17/2018 16:30	SMPL
CCB018	CCB018	-35	-0.0372 -		ISM02.4	PPB	10/17/2018 16:32	SMPL
J5387-20	MYAKN6	849	0.1745 -		ISM02.4	PPB	10/17/2018 16:42	SMPL
J5387-21	MYAKN6D	876	0.1809 -		ISM02.4	PPB	10/17/2018 16:44	SMPL
J5387-22	MYAKN6S	12934	3.0681 -		ISM02.4	PPB	10/17/2018 16:47	SMPL
PB113904BL	PBS004	-37	-0.0377 -		ISM02.4	PPB	10/17/2018 16:49	SMPL
J5388-01	MYAKN7	590	0.1124 -		ISM02.4	PPB	10/17/2018 16:51	SMPL
J5388-02	MYAKN8	466	0.0828 -		ISM02.4	PPB	10/17/2018 16:54	SMPL
J5388-03	MYAKP1	162	0.01 -		ISM02.4	PPB	10/17/2018 16:56	SMPL
J5388-04	MYAKP2	599	0.1146 -		ISM02.4	PPB	10/17/2018 16:58	SMPL
J5388-05	MYAKY9	120	-0.0001 -		ISM02.4	PPB	10/17/2018 17:00	SMPL
J5388-06	MYAJX4	552	0.1033 -		ISM02.4	PPB	10/17/2018 17:03	SMPL
J5388-07	MYAKN3	956	0.2001 -		ISM02.4	PPB	10/17/2018 17:05	SMPL
J5388-08	MYAKN9	1113	0.2377 -		ISM02.4	PPB	10/17/2018 17:07	SMPL
J5388-09	MYAJW6	1376	0.3006 -		ISM02.4	PPB	10/17/2018 17:10	SMPL
J5388-10	MYAJW7	1063	0.2257 -		ISM02.4	PPB	10/17/2018 17:12	SMPL

INSTRUMENT ID : CV1

J5388-11	MYAJW8	684	0.1349 -	ISM02.4	PPB	10/17/2018 17:14 SMPL
J5388-12	MYAJW9	900	0.1867 -	ISM02.4	PPB	10/17/2018 17:16 SMPL
J5388-13	MYAJX0	928	0.1934 -	ISM02.4	PPB	10/17/2018 17:19 SMPL
J5388-14	MYAJX1	611	0.1175 -	ISM02.4	PPB	10/17/2018 17:21 SMPL
J5388-15	MYAJX2	649	0.1266 -	ISM02.4	PPB	10/17/2018 17:23 SMPL
J5388-16	MYAJX3	514	0.0942 -	ISM02.4	PPB	10/17/2018 17:25 SMPL
CCV019	CCV019	21106	5.0248 -	ISM02.4	PPB	10/17/2018 17:28 SMPL
CCB019	CCB019	-84	-0.0489 -	ISM02.4	PPB	10/17/2018 17:30 SMPL
J5388-17	MYAJY1	638	0.1239 -	ISM02.4	PPB	10/17/2018 17:32 SMPL
J5388-18	MYAJY2	184	0.0152 -	ISM02.4	PPB	10/17/2018 17:35 SMPL
J5388-19	MYAJY3	156	0.0085 -	ISM02.4	PPB	10/17/2018 17:37 SMPL
J5388-20	MYAJY4	101	-0.0046 -	ISM02.4	PPB	10/17/2018 17:39 SMPL
J5388-21	MYAJY4D	99	-0.0051 -	ISM02.4	PPB	10/17/2018 17:41 SMPL
J5388-22	MYAJY4S	12419	2.9448 -	ISM02.4	PPB	10/17/2018 17:44 SMPL
PB113905BL	PBS004	-6	-0.0303 -	ISM02.4	PPB	10/17/2018 17:46 SMPL
J5389-01	MYAJY5	149	0.0068 -	ISM02.4	PPB	10/17/2018 17:48 SMPL
J5389-02	MYAJY6	111	-0.0023 -	ISM02.4	PPB	10/17/2018 17:51 SMPL
J5389-03	MYAJY7	138	0.0042 -	ISM02.4	PPB	10/17/2018 17:53 SMPL
J5389-04	MYAJY8	198	0.0186 -	ISM02.4	PPB	10/17/2018 17:55 SMPL
J5389-05	MYAJY9	202	0.0195 -	ISM02.4	PPB	10/17/2018 17:57 SMPL
J5389-06	MYAJZ0	183	0.015 -	ISM02.4	PPB	10/17/2018 18:00 SMPL
J5389-07	MYAJZ6	708	0.1407 -	ISM02.4	PPB	10/17/2018 18:02 SMPL
J5389-08	MYAJZ7	1494	0.3289 -	ISM02.4	PPB	10/17/2018 18:04 SMPL
J5389-09	MYAJZ8	2805	0.6428 -	ISM02.4	PPB	10/17/2018 18:07 SMPL
J5389-10	MYAJZ9	526	0.0971 -	ISM02.4	PPB	10/17/2018 18:09 SMPL
J5389-11	MYAJZ9D	453	0.0796 -	ISM02.4	PPB	10/17/2018 18:11 SMPL
J5389-12	MYAJZ9S	12579	2.9831 -	ISM02.4	PPB	10/17/2018 18:13 SMPL
CCV020	CCV020	21025	5.0054 -	ISM02.4	PPB	10/17/2018 18:16 SMPL
CCB020	CCB020	-108	-0.0547 -	ISM02.4	PPB	10/17/2018 18:18 SMPL