



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB98665

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	10/17/2018	15:45
2	CAL2	1	Water			22.4	7.00	10/17/2018	15:47
3	CAL3	1	Water			22.3	10.04	10/17/2018	15:49
4	ICV	1	Water			22.2	7.01	10/17/2018	15:51
5	CCV1	1	Water			22.5	2.00	10/17/2018	15:53
6	J5569-01	1	Water			21.3	6.35	10/17/2018	15:55
7	J5569-01DUP	1	Water			21.4	6.36	10/17/2018	15:57
8	J5569-02	1	Water			21.2	6.33	10/17/2018	15:59
9	J5569-05	1	Water			21.6	7.01	10/17/2018	16:01
10	J5569-06	1	Water			21.3	7.74	10/17/2018	16:03
11	J5569-07	1	Water			22.0	7.54	10/17/2018	16:05
12	J5569-08	1	Water			21.6	7.41	10/17/2018	16:07
13	J5569-09	1	Water			21.6	7.10	10/17/2018	16:09
14	J5569-12	1	Water			21.8	7.34	10/17/2018	16:11
15	J5569-13	1	Water			21.6	6.41	10/17/2018	16:13
16	CCV2	1	Water			22.3	12.00	10/17/2018	16:15

WORKLIST(Hardcopy Internal Chain)

LB 98665

WorkList Name : pH101718-3

WorkList ID : 118143

Date : 10/17/2018 3:20:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/16/2018	Water	J5569-01	K pH	Cool 4 deg C	CHAZ02	D31	RLF-GW-7S	10/16/2018	SM 4500-H B
10/16/2018	Water	J5569-02	K pH	Cool 4 deg C	CHAZ02	D31	RLF-GW-9S	10/16/2018	SM 4500-H B
10/16/2018	Water	J5569-05	K pH	Cool 4 deg C	CHAZ02	D31	RLF-GW-FD-01	10/16/2018	SM 4500-H B
10/16/2018	Water	J5569-06	K pH	Cool 4 deg C	CHAZ02	D31	RLF-SW-1	10/16/2018	SM 4500-H B
10/16/2018	Water	J5569-07	K pH	Cool 4 deg C	CHAZ02	D31	RLF-SW-3	10/16/2018	SM 4500-H B
10/16/2018	Water	J5569-08	K pH	Cool 4 deg C	CHAZ02	D31	RLF-SW-4	10/16/2018	SM 4500-H B
10/16/2018	Water	J5569-09	K pH	Cool 4 deg C	CHAZ02	D31	RLF-SW-5	10/16/2018	SM 4500-H B
10/16/2018	Water	J5569-12	K pH	Cool 4 deg C	CHAZ02	D31	RLF-SW-8	10/16/2018	SM 4500-H B
10/16/2018	Water	J5569-13	K pH	Cool 4 deg C	CHAZ02	D31	RLF-SW-FD-01	10/16/2018	SM 4500-H B

Date/Time 10/17/18 3:25
 Received by: SP
 Relinquished by: CP

Date/Time 10/17/18 4:20
 Received by: CP
 Relinquished by: SP