


7471B

LB 98674
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42680



A= 0.0000e+000

B= 2.3628e-004

C= 7.6807e-003

Rho= 0.9997165

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	140	0.000	140				
0.2	0.200	0.189	-0.011	767	0.0 %	767				
2.5	2.500	2.471	-0.029	10427	0.0 %	10427				
5.0	5.000	5.078	0.078	21459	0.0 %	21459				
7.5	7.500	7.329	-0.171	30987	0.0 %	30987				
10.0	10.000	10.092	0.092	42680	0.0 %	42680				

LB98674 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Oct 2018 18:38:43

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	140	-	0.00007471B	PPB		17 Oct 2018 18:49:32	S	Std
0.2 - 1	50.2	767	-	0.20007471B	PPB		17 Oct 2018 18:51:50	S	Std
2.5 - 1	52.5	10427	-	2.50007471B	PPB		17 Oct 2018 18:54:08	S	Std
5.0 - 1	55	21459	-	5.00007471B	PPB		17 Oct 2018 18:56:25	S	Std
7.5 - 1	57.5	30987	-	7.50007471B	PPB		17 Oct 2018 18:58:43	S	Std
10.0 - 1	510	42680	-	10.00007471B	PPB		17 Oct 2018 19:00:59	S	Std
ICV27 - 1	ICV27	16302	3.8594	-7471B	PPB		17 Oct 2018 19:19:24	U	SMPL
ICB27 - 1	ICB27	68	0.0237	-7471B	PPB		17 Oct 2018 19:21:40	U	SMPL
CCV21 - 1	CCV21	21658	5.1249	-7471B	PPB		17 Oct 2018 19:23:55	U	SMPL
CCB21 - 1	CCB21	29	0.0145	-7471B	PPB		17 Oct 2018 19:26:11	U	SMPL
CRA - 1	CRA	736	0.1816	-7471B	PPB		17 Oct 2018 19:28:28	U	SMPL
HighStd - 1	HighStd	42184	9.9747	-7471B	PPB		17 Oct 2018 19:30:44	U	SMPL
ChkStd - 1	ChkStd	29306	6.9320	-7471B	PPB		17 Oct 2018 19:33:00	U	SMPL
PB113912BL - 1	PBS	-196	-0.0386	-7471B	PPB		17 Oct 2018 19:35:19	U	SMPL
PB113912BS - 1	LCSS	15135	3.5837	-7471B	PPB		17 Oct 2018 19:37:37	U	SMPL
J5200-09 - 1	OR-01-101618-A	1780	0.4283	-7471B	PPB		17 Oct 2018 19:39:54	U	SMPL
J5200-11 - 1	OR-01-101618-B	2886	0.6896	-7471B	PPB		17 Oct 2018 19:42:11	U	SMPL
J5518-01 - 1	JCTS	2820	0.6740	-7471B	PPB		17 Oct 2018 19:44:28	U	SMPL
J5518-02 - 1	SODUS-TS	2820	0.6740	-7471B	PPB		17 Oct 2018 19:46:46	U	SMPL
J5518-03 - 1	GSTS	2341	0.5608	-7471B	PPB		17 Oct 2018 19:49:02	U	SMPL
CCV22 - 1	CCV22	21855	5.1715	-7471B	PPB		17 Oct 2018 19:51:18	U	SMPL
CCB22 - 1	CCB22	67	0.0235	-7471B	PPB		17 Oct 2018 19:53:34	U	SMPL
J5527-01 - 1	SP-STONE	206	0.0564	-7471B	PPB		17 Oct 2018 19:55:51	U	SMPL
J5529-01 - 1	TS	9203	2.1821	-7471B	PPB		17 Oct 2018 19:58:07	U	SMPL
J5535-01 - 1	BB-DRUM	2951	0.7049	-7471B	PPB		17 Oct 2018 20:00:23	U	SMPL
J5536-01 - 1	351-SOIL	56113	13.2658	-7471B	PPB		17 Oct 2018 20:02:39	U	SMPL
J5539-01 - 1	SB-6-0.3-2FT	889	0.2177	-7471B	PPB		17 Oct 2018 20:04:57	U	SMPL
J5539-02 - 1	SB-8-3-5FT	3177	0.7583	-7471B	PPB		17 Oct 2018 20:07:16	U	SMPL
J5539-03 - 1	DUP	1988	0.4774	-7471B	PPB		17 Oct 2018 20:09:34	U	SMPL
J5539-07 - 1	SB-10-5-7FT	968	0.2364	-7471B	PPB		17 Oct 2018 20:16:41	U	SMPL
J5539-07DUP - 1	SB-10-5-7FTDUP	945	0.2310	-7471B	PPB		17 Oct 2018 20:18:58	U	SMPL
CCV23 - 1	CCV23	21998	5.2053	-7471B	PPB		17 Oct 2018 20:21:15	U	SMPL
CCB23 - 1	CCB23	74	0.0252	-7471B	PPB		17 Oct 2018 20:23:32	U	SMPL
J5539-08 - 1	J5539-07MS	16569	3.9225	-7471B	PPB		17 Oct 2018 20:25:48	U	SMPL
J5539-09 - 1	J5539-07MSD	16288	3.8561	-7471B	PPB		17 Oct 2018 20:28:04	U	SMPL
J5539-07LX5 - 1		213	0.0580	-7471B	PPB		17 Oct 2018 20:30:20	U	SMPL
J5539-07A - 1		17680	4.1850	-7471B	PPB		17 Oct 2018 20:32:38	U	SMPL
J5536-01DLX2 - 1		28624	6.7708	-7471B	PPB		17 Oct 2018 20:34:54	U	SMPL
CCV23 - 1	CCV23	21734	5.1429	-7471B	PPB		17 Oct 2018 20:37:11	U	SMPL
CCB23 - 1	CCB23	18	0.0119	-7471B	PPB		17 Oct 2018 20:39:29	U	SMPL

10/29/18