



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98679

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.5	4.01	10/17/2018	17:10
2	CAL2	1	Water			22.7	7.00	11/17/2018	17:12
3	CAL3	1	Water			22.8	10.04	10/17/2018	17:14
4	ICV	1	Water			22.6	7.01	10/17/2018	17:16
5	CCV1	1	Water			22.5	2.00	10/17/2018	17:18
6	J5202-02	1	Solid	20.02	20	22.8	6.57	10/17/2018	17:20
7	J5202-04	1	Solid	20.04	20	22.6	7.03	10/17/2018	17:22
8	J5202-06	1	Solid	20.02	20	22.5	6.48	10/17/2018	17:24
9	J5204-06	1	Solid	20.02	20	21.8	8.44	10/17/2018	17:26
10	J5204-08	1	Solid	20.02	20	22.3	7.52	10/17/2018	17:28
11	J5549-01	1	Solid	20.03	20	20.3	8.30	10/17/2018	17:30
12	J5549-01DUP	1	Solid	20.04	20	20.4	8.31	10/17/2018	17:32
13	J5549-02	1	Solid	20.02	20	20.4	8.41	10/17/2018	17:34
14	J5579-04	1	Solid	20.03	20	22.4	8.18	10/17/2018	17:36
15	CCV2	1	Water			22.6	12.00	10/17/2018	17:38

WORKLIST(Hardcopy Internal Chain)

UB 98679

WorkList Name : corrosivity101718

WorkList ID : 118147

Date : 10/17/2018 3:55:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/17/2018	Solid	J5202-02	Corrosivity	Cool 4 deg C	PSEG05	P63	PL-03-101718-A	10/17/2018	9045D
10/17/2018	Solid	J5202-04	Corrosivity	Cool 4 deg C	PSEG05	P63	PL-03-101718-B	10/17/2018	9045D
10/17/2018	Solid	J5202-06	Corrosivity	Cool 4 deg C	PSEG05	P63	PL-03-101718-C	10/17/2018	9045D
10/17/2018	Solid	J5204-06	Corrosivity	Cool 4 deg C	PSEG05		SU-04-101718-A	10/17/2018	9045D
10/17/2018	Solid	J5204-08	Corrosivity	Cool 4 deg C	PSEG05		SU-04-101718-B	10/17/2018	9045D
10/15/2018	Solid	J5549-01	Corrosivity	Cool 4 deg C	USAR03	D42	CB-1018-S-TCLP-GRID-45	10/15/2018	9045D
10/15/2018	Solid	J5549-02	Corrosivity	Cool 4 deg C	USAR03	D42	CB-1018-S-TCLP-GRID-50-2	10/15/2018	9045D
10/17/2018	Solid	J5579-04	Corrosivity	Cool 4 deg C	PSEG01	P52	7-1-ROLLOFFS	10/17/2018	9045D

Date/Time 10/17/18 5:00
 Received by: SP
 Relinquished by: CP

Date/Time 10/17/18 5:40
 Received by: CP
 Relinquished by: LSP